

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Homeowner **SDG No.:** Q1410
Contract: HOME01 **Lab Code:** CHEM **Case No.:** Q1410 **SAS No.:** Q1410
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	17.0	2.0	850	-18	22	03/06/2025	11:51	LB134928
	Lead	6.40			-12	12	03/06/2025	11:51	LB134928
ICSAB01	Copper	537	511	105	434	588	03/06/2025	12:04	LB134928
	Lead	55.0	49.0	112	37	61	03/06/2025	12:04	LB134928
ICSA	Copper	29.0	2.0	1450	-18	22	03/06/2025	12:10	LB134928
	Lead	40.5			-12	12	03/06/2025	12:10	LB134928
ICSAB	Copper	574	511	112	434	588	03/06/2025	12:14	LB134928
	Lead	56.6	49.0	116	37	61	03/06/2025	12:14	LB134928