

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Homeowner **SDG No.:** Q1410
Contract: HOME01 **Lab Code:** CHEM **Case No.:** Q1410 **SAS No.:** Q1410
Matrix: Water **Level:** LOW **Client ID:** KITCHEN-TAPA
Sample ID: Q1410-01 **Spiked ID:** Q1410-01A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	85 - 115	665		571		150	63		P