

7470A


65134804

INSTRUMENT ID: CV1

Linear

μ Abs.:

42360



A= 0.0000e+000

B= 2.3390e-004

C= -7.6175e-002

Rho= 0.9989125

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/d
0.0	0.000	-0.033	-0.033	186	0.000	186	0				0/0
0.2	0.200	0.173	-0.027	1064	0.0 %	1064					-14
2.5	2.500	2.361	-0.139	10421	0.0 %	10421					-6
5.0	5.000	5.358	0.358	23233	0.0 %	23233					7
7.5	7.500	7.509	0.009	32429	0.0 %	32429					0
10.0	10.000	9.832	-0.168	42360	0.0 %	42360					-2

LB134804 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Feb 2025 08:16:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	5.2	186	-	0.0000	7470A	PPB	26 Feb 2025 08:21:01	S	Std
0.2 - 1	5.2	1064	-	0.2000	7470A	PPB	26 Feb 2025 08:23:19	S	Std
2.5 - 1	5.2	10421	-	2.5000	7470A	PPB	26 Feb 2025 08:25:36	S	Std
5.0 - 1	5.2	23233	-	5.0000	7470A	PPB	26 Feb 2025 08:27:53	S	Std
7.5 - 1	5.2	32429	-	7.5000	7470A	PPB	26 Feb 2025 08:30:13	S	Std
10.0 - 1	5.2	42360	-	10.0000	7470A	PPB	26 Feb 2025 08:32:32	S	Std
ICV17 - 1	ICV17	18155	4.1703	-	7470A	PPB	26 Feb 2025 08:35:26	U	SMPL
ICB17 - 1	ICB17	213	-0.0264	-	7470A	PPB	26 Feb 2025 08:40:18	U	SMPL
CCV18 - 1	CCV18	23251	5.3622	-	7470A	PPB	26 Feb 2025 08:42:32	U	SMPL
CCB18 - 1	CCB18	-77	-0.0942	-	7470A	PPB	26 Feb 2025 08:44:47	U	SMPL
CRA - 1	CRA	1025	0.1636	-	7470A	PPB	26 Feb 2025 08:52:34	U	SMPL
HighStd - 1	HighStd	41745	9.6879	-	7470A	PPB	26 Feb 2025 08:54:48	U	SMPL
ChkStd - 1	ChkStd	29838	6.9029	-	7470A	PPB	26 Feb 2025 08:57:03	U	SMPL
PB166869BL - 1	PBW	236	-0.0210	-	7470A	PPB	26 Feb 2025 09:02:17	U	SMPL
PB166869BS - 1	LCSW	17330	3.9773	-	7470A	PPB	26 Feb 2025 09:04:33	U	SMPL
Q1409-01 - 1	SURFACE-WATER-1	-287	-0.1433	-	7470A	PPB	26 Feb 2025 09:06:49	U	SMPL
Q1409-01DUP - 1	SURFACE-WATER-1DUP	-54	-0.0888	-	7470A	PPB	26 Feb 2025 09:09:07	U	SMPL
Q1409-01MS - 1	SURFACE-WATER-1MS	16238	3.7219	-	7470A	PPB	26 Feb 2025 09:11:24	U	SMPL
Q1409-01MSD - 1	SURFACE-WATER-1MSD	16315	3.7399	-	7470A	PPB	26 Feb 2025 09:13:41	U	SMPL
Q1415-05 - 1	FB02222025	-349	-0.1578	-	7470A	PPB	26 Feb 2025 09:15:59	U	SMPL
CCV19 - 1	CCV19	23051	5.3154	-	7470A	PPB	26 Feb 2025 09:18:16	U	SMPL
CCB19 - 1	CCB19	-97	-0.0989	-	7470A	PPB	26 Feb 2025 09:20:32	U	SMPL
Q1415-06 - 1	B-173-GW01	5024	1.0989	-	7470A	PPB	26 Feb 2025 09:22:49	U	SMPL
Q1409-01LX5 - 1		-647	-0.2275	-	7470A	PPB	26 Feb 2025 09:25:05	U	SMPL
Q1409-01A - 1		17073	3.9172	-	7470A	PPB	26 Feb 2025 09:31:36	U	SMPL
CCV20 - 1	CCV20	20771	4.7821	-	7470A	PPB	26 Feb 2025 09:33:51	U	SMPL
CCB20 - 1	CCB20	-91	-0.0975	-	7470A	PPB	26 Feb 2025 09:36:09	U	SMPL