

Continuing Calibration Report for SequenceID : FC022525AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068295.D

Aliphatic C9-C12	7540145.000	60.000	3.105	6.371	125669.083	141035.502	10.895
Aliphatic C12-C16	4890516.000	40.000	6.372	9.756	122262.900	136654.515	10.531
Aliphatic C16-C21	7157687.000	60.000	9.757	13.110	119294.783	131630.895	9.372
Aliphatic C21-C28	9437216.000	80.000	13.111	16.764	117965.200	122691.322	3.852
Aliphatic C28-C40	12705875.000	120.000	16.765	21.571	105882.292	102846.462	-2.952
Aliphatic EPH	41731439.000	360.000	3.105	21.571	115920.664	122175.127	5.119