

Continuing Calibration Report for SequenceID : FC022825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068321.D

Aliphatic C9-C12	8795568.000	60.000	3.104	6.369	146592.800	141035.502	-3.940
Aliphatic C12-C16	5630378.000	40.000	6.370	9.754	140759.450	136654.515	-3.004
Aliphatic C16-C21	8123361.000	60.000	9.755	13.107	135389.350	131630.895	-2.855
Aliphatic C21-C28	10243894.000	80.000	13.108	16.761	128048.675	122691.322	-4.367
Aliphatic C28-C40	13204528.000	120.000	16.762	21.565	110037.733	102846.462	-6.992
Aliphatic EPH	45997729.000	360.000	3.104	21.565	127771.469	122175.127	-4.581