

Continuing Calibration Report for SequenceID : FC022825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068331.D

Aliphatic C9-C12	8595776.000	60.000	3.104	6.369	143262.933	141035.502	-1.579
Aliphatic C12-C16	5515542.000	40.000	6.370	9.754	137888.550	136654.515	-0.903
Aliphatic C16-C21	7931958.000	60.000	9.755	13.107	132199.300	131630.895	-0.432
Aliphatic C21-C28	9960510.000	80.000	13.108	16.761	124506.375	122691.322	-1.479
Aliphatic C28-C40	13030922.000	120.000	16.762	21.565	108591.017	102846.462	-5.586
Aliphatic EPH	45034708.000	360.000	3.104	21.565	125096.411	122175.127	-2.391