

Continuing Calibration Report for SequenceID : FD022525AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049100.D

Aromatic C16-C21	12146265.000	80.000	8.397	12.660	151828.313	153359.309	0.998
Aromatic C21-C36	26130774.000	180.000	12.661	18.070	145170.967	143114.753	-1.437
Aromatic EPH	54757193.000	360.000	4.075	18.070	152103.314	150128.823	-1.315
Aromatic C10-C12	6565172.000	40.000	4.075	5.793	164129.300	159020.606	-3.213
Aromatic C12-C16	9914982.000	60.000	5.794	8.396	165249.700	160935.866	-2.680