

Continuing Calibration Report for SequenceID : FD022525AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049107.D

Aromatic C16-C21	12153015.000	80.000	8.397	12.660	151912.688	153359.309	0.943
Aromatic C21-C36	25896026.000	180.000	12.661	18.070	143866.811	143114.753	-0.525
Aromatic EPH	54443876.000	360.000	4.075	18.070	151232.989	150128.823	-0.735
Aromatic C10-C12	6541813.000	40.000	4.075	5.793	163545.325	159020.606	-2.845
Aromatic C12-C16	9853022.000	60.000	5.794	8.396	164217.033	160935.866	-2.039