

SOLID EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1415 **SAS No :** Q1415 **SDG No:** Q1415
Sample No : Q1421-03 **Datafile:** FC068335.D
Client ID : P001-CLAY-CF01-01MSD

COMPOUND	SPIKE ADDED mg/kg	SAMPLE CONCENTRATION mg/kg	MS/MSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS QC Limit Of RPD
Aliphatic C9-C12	14.2	0	9.71	68		0.15 (40-140) 25
Aliphatic C12-C16	9.5	0.50	9.10	91		0.44 (40-140) 25
Aliphatic C16-C21	14.2	0.43	14.7	100		0.1 (40-140) 25
Aliphatic C21-C28	18.9	0	19.6	103		0.98 (40-140) 25
Aliphatic C28-C40	28.4	2.80	34.6	111		0.9 (40-140) 25