

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1415 **SAS No :** Q1415 **SDG No:** Q1415
Sample No : PB166863BS **Datafile:** FC068297.D
Client ID : PB166863BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aliphatic C9-C12	300	172	57		(40-140)
Aliphatic C12-C16	200	158	79		(40-140)
Aliphatic C16-C21	300	268	89		(40-140)
Aliphatic C21-C28	400	366	92		(40-140)
Aliphatic C28-C40	600	602	100		(40-140)