

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1415 **SAS No :** Q1415 **SDG No:** Q1415
Sample No : PB166863BSD **Datafile:** FC068298.D
Client ID : PB166863BSD

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	300	175	58		1.8 (40-140)	50
Aliphatic C12-C16	200	160	80		1.3 (40-140)	50
Aliphatic C16-C21	300	268	89		0.03 (40-140)	50
Aliphatic C21-C28	400	364	91		0.548 (40-140)	50
Aliphatic C28-C40	600	602	100		0.33 (40-140)	50