

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Portal Partners Tri-Venture
Lab Code: CHEM **Cas No:** Q1415 **SAS No :** Q1415 **SDG No:** Q1415
Sample No : PB166863BS **Datafile:** FD049102.D
Client ID : PB166863BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aromatic C10-C12	200	165	82		(40-140)
Aromatic C12-C16	300	410	136		(40-140)
Aromatic C16-C21	500	583	116		(40-140)
Aromatic C21-C36	800	905	113		(40-140)