

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1415 SAS No.: Q1415 SDG NO.: Q1415

Continuing Calib Date: 02/25/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 09:59 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.68	5.68	5.58	5.78	0.00
Aroclor-1242-2 (2)	5.70	5.70	5.60	5.80	0.00
Aroclor-1242-3 (3)	5.76	5.77	5.67	5.87	0.01
Aroclor-1242-4 (4)	5.86	5.86	5.76	5.96	0.00
Aroclor-1242-5 (5)	6.59	6.59	6.49	6.69	0.00
Tetrachloro-m-xylene	4.53	4.53	4.43	4.63	0.00
Decachlorobiphenyl	10.26	10.26	10.16	10.36	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1415 SAS No.: Q1415 SDG NO.: Q1415

Continuing Calib Date: 02/25/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 09:59 Initial Calibration Time(s): 14:59 22:17

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.92	4.92	4.82	5.02	0.00
Aroclor-1242-2	(2)	4.94	4.94	4.84	5.04	0.00
Aroclor-1242-3	(3)	5.12	5.12	5.02	5.22	0.00
Aroclor-1242-4	(4)	5.20	5.20	5.10	5.30	0.00
Aroclor-1242-5	(5)	5.73	5.73	5.63	5.83	0.01
Tetrachloro-m-xylene		3.83	3.83	3.73	3.93	0.00
Decachlorobiphenyl		8.89	8.89	8.79	8.99	0.00