

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1415 SAS No.: Q1415 SDG NO.: Q1415

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 02/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP070045.D Time Analyzed: 15:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.682	5.581	5.781	493.070	500.000	-1.4
Aroclor-1016-2	5.703	5.602	5.802	488.280	500.000	-2.3
Aroclor-1016-3	5.766	5.665	5.865	485.740	500.000	-2.9
Aroclor-1016-4	5.863	5.762	5.962	479.880	500.000	-4.0
Aroclor-1016-5	6.156	6.056	6.256	494.780	500.000	-1.0
Aroclor-1260-1	7.276	7.175	7.375	470.130	500.000	-6.0
Aroclor-1260-2	7.529	7.429	7.629	457.880	500.000	-8.4
Aroclor-1260-3	7.888	7.788	7.988	484.630	500.000	-3.1
Aroclor-1260-4	8.113	8.012	8.212	477.180	500.000	-4.6
Aroclor-1260-5	8.433	8.333	8.533	494.540	500.000	-1.1
Decachlorobiphenyl	10.258	10.156	10.356	49.540	50.000	-0.9
Tetrachloro-m-xylene	4.528	4.427	4.627	50.690	50.000	1.4

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1415 SAS No.: Q1415 SDG NO.: Q1415

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 02/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP070045.D Time Analyzed: 15:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.922	4.820	5.020	474.050	500.000	-5.2
Aroclor-1016-2	4.941	4.839	5.039	475.050	500.000	-5.0
Aroclor-1016-3	5.118	5.017	5.217	482.530	500.000	-3.5
Aroclor-1016-4	5.160	5.058	5.258	484.710	500.000	-3.1
Aroclor-1016-5	5.375	5.273	5.473	481.080	500.000	-3.8
Aroclor-1260-1	6.413	6.312	6.512	468.300	500.000	-6.3
Aroclor-1260-2	6.601	6.500	6.700	476.360	500.000	-4.7
Aroclor-1260-3	6.755	6.654	6.854	428.530	500.000	-14.3
Aroclor-1260-4	7.228	7.126	7.326	444.280	500.000	-11.1
Aroclor-1260-5	7.468	7.367	7.567	467.420	500.000	-6.5
Decachlorobiphenyl	8.890	8.788	8.988	46.920	50.000	-6.2
Tetrachloro-m-xylene	3.832	3.730	3.930	48.180	50.000	-3.6