

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1415 SAS No.: Q1415 SDG NO.: Q1415

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL03 Date Analyzed: 02/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP070057.D Time Analyzed: 19:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.681	5.581	5.781	504.510	500.000	0.9
Aroclor-1016-2	5.702	5.602	5.802	518.780	500.000	3.8
Aroclor-1016-3	5.764	5.665	5.865	596.270	500.000	19.3
Aroclor-1016-4	5.862	5.762	5.962	533.560	500.000	6.7
Aroclor-1016-5	6.154	6.056	6.256	597.690	500.000	19.5
Aroclor-1260-1	7.274	7.175	7.375	546.230	500.000	9.2
Aroclor-1260-2	7.528	7.429	7.629	485.180	500.000	-3.0
Aroclor-1260-3	7.886	7.788	7.988	518.870	500.000	3.8
Aroclor-1260-4	8.111	8.012	8.212	508.370	500.000	1.7
Aroclor-1260-5	8.432	8.333	8.533	517.950	500.000	3.6
Decachlorobiphenyl	10.255	10.156	10.356	51.200	50.000	2.4
Tetrachloro-m-xylene	4.528	4.427	4.627	53.380	50.000	6.8

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: Q1415 SAS No.: Q1415 SDG NO.: Q1415

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL03 Date Analyzed: 02/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP070057.D Time Analyzed: 19:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.920	4.820	5.020	505.210	500.000	1.0
Aroclor-1016-2	4.939	4.839	5.039	502.910	500.000	0.6
Aroclor-1016-3	5.117	5.017	5.217	516.640	500.000	3.3
Aroclor-1016-4	5.158	5.058	5.258	509.820	500.000	2.0
Aroclor-1016-5	5.373	5.273	5.473	534.740	500.000	6.9
Aroclor-1260-1	6.412	6.312	6.512	491.190	500.000	-1.8
Aroclor-1260-2	6.599	6.500	6.700	491.950	500.000	-1.6
Aroclor-1260-3	6.753	6.654	6.854	476.690	500.000	-4.7
Aroclor-1260-4	7.226	7.126	7.326	539.140	500.000	7.8
Aroclor-1260-5	7.467	7.367	7.567	529.100	500.000	5.8
Decachlorobiphenyl	8.888	8.788	8.988	53.790	50.000	7.6
Tetrachloro-m-xylene	3.831	3.730	3.930	50.160	50.000	0.3