

## Report of Analysis

Client:	Portal Partners Tri-Venture		Date Collected:	02/24/25	
Project:	Amtrak Sawtooth Bridges 2025		Date Received:	02/24/25	
Client Sample ID:	P001-CLAY-CF01-01MSD		SDG No.:	Q1415	
Lab Sample ID:	Q1421-03MSD		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	70.3	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP070049.D	1	02/25/25 09:02	02/25/25 16:32	PB166854

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
<b>TARGETS</b>						
12674-11-2	Aroclor-1016	249		4.80	24.2	ug/kg
11104-28-2	Aroclor-1221	9.10	U	9.10	24.2	ug/kg
11141-16-5	Aroclor-1232	4.80	U	4.80	24.2	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	24.2	ug/kg
12672-29-6	Aroclor-1248	11.2	U	11.2	24.2	ug/kg
11097-69-1	Aroclor-1254	3.90	U	3.90	24.2	ug/kg
37324-23-5	Aroclor-1262	6.50	U	6.50	24.2	ug/kg
11100-14-4	Aroclor-1268	4.90	U	4.90	24.2	ug/kg
11096-82-5	Aroclor-1260	226		4.10	24.2	ug/kg
<b>SURROGATES</b>						
877-09-8	Tetrachloro-m-xylene	28.0		30 (32) - 150 (144)	140%	SPK: 20
2051-24-3	Decachlorobiphenyl	26.0		30 (32) - 150 (175)	130%	SPK: 20

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit