



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1415

Client: Portal Partners Tri-Venture

Analytical Method: 8082A Datafile : PP070054.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB166856BSD	AR1016	5	5.00	ug/L	100	2			40 (61)	140 (112)	20 (20)
	AR1260	5	4.70	ug/L	94	2			40 (66)	140 (113)	20 (20)