

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: Q1415 SDG NO.: Q1415

Lab File ID: BF141774.D

DFTPP Injection Date: 02/26/2025

Instrument ID: BNA_F

DFTPP Injection Time: 10:16

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	37
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	33.5
70	Less than 2.0% of mass 69	0.1 (0.4) 1
127	10.0 - 80.0% of mass 198	47.6
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	27.2
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	14
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	16.7 (18.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF141775.D	02/26/2025	10:45
PB166853BL	PB166853BL	BF141776.D	02/26/2025	11:49
PB166853BS	PB166853BS	BF141777.D	02/26/2025	12:19
PB166875BL	PB166875BL	BF141780.D	02/26/2025	13:55
PB166875BS	PB166875BS	BF141781.D	02/26/2025	14:24
PB166875BSD	PB166875BSD	BF141782.D	02/26/2025	14:54
B-173-GW01	Q1415-06	BF141787.D	02/26/2025	17:27
FB02222025	Q1415-05	BF141788.D	02/26/2025	17:56