

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QCBatch ID # VN021825

Review By	John Carlone	Review On	2/19/2025 9:44:55 AM
Supervise By	Maresh Dadoda	Supervise On	2/19/2025 1:34:51 PM
SubDirectory	VN021825	HP Acquire Method	HP Processing Method 82N021825W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133068		
Initial Calibration Stds	VP133070,VP133071,VP133072,VP133073,VP133074,VP133075		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP133076		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN085771.D	18 Feb 2025 10:35	JC\MD	Ok
2	VSTDICC100	VN085772.D	18 Feb 2025 11:09	JC\MD	Ok,M
3	VSTDICCC050	VN085773.D	18 Feb 2025 11:32	JC\MD	Ok,M
4	VSTDICC020	VN085774.D	18 Feb 2025 11:56	JC\MD	Not Ok
5	VSTDICC010	VN085775.D	18 Feb 2025 12:20	JC\MD	Ok,M
6	VSTDICC005	VN085776.D	18 Feb 2025 12:43	JC\MD	Ok,M
7	VSTDICC001	VN085777.D	18 Feb 2025 13:07	JC\MD	Ok,M
8	IBLK	VN085778.D	18 Feb 2025 13:54	JC\MD	Ok
9	VSTDICC020	VN085779.D	18 Feb 2025 14:18	JC\MD	Ok,M
10	VSTDICV050	VN085780.D	18 Feb 2025 16:28	JC\MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN022425

Review By	John Carlone	Review On	2/25/2025 9:55:02 AM
Supervise By	Maresh Dadoda	Supervise On	2/25/2025 2:29:22 PM
SubDirectory	VN022425	HP Acquire Method	HP Processing Method 82N021825W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133126		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133127,VP133128		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN085828.D	24 Feb 2025 10:50	JC\MD	Ok
2	VSTDCCC050	VN085829.D	24 Feb 2025 11:37	JC\MD	Ok,M
3	VN0224MBL01	VN085830.D	24 Feb 2025 12:15	JC\MD	Ok
4	VN0224WBL01	VN085831.D	24 Feb 2025 12:39	JC\MD	Ok
5	VN0224WBS01	VN085832.D	24 Feb 2025 13:03	JC\MD	Ok,M
6	VN0224WBSD01	VN085833.D	24 Feb 2025 13:36	JC\MD	Ok,M
7	Q1405-02DL	VN085834.D	24 Feb 2025 14:00	JC\MD	Ok
8	Q1405-03	VN085835.D	24 Feb 2025 14:24	JC\MD	Ok
9	Q1405-04	VN085836.D	24 Feb 2025 14:49	JC\MD	Ok,M
10	Q1405-05	VN085837.D	24 Feb 2025 15:13	JC\MD	Ok
11	Q1405-11	VN085838.D	24 Feb 2025 15:37	JC\MD	Ok
12	Q1415-05	VN085839.D	24 Feb 2025 16:01	JC\MD	ReRun
13	Q1405-07	VN085840.D	24 Feb 2025 16:26	JC\MD	Ok
14	Q1405-08	VN085841.D	24 Feb 2025 16:50	JC\MD	Ok
15	Q1405-09	VN085842.D	24 Feb 2025 17:14	JC\MD	Ok
16	Q1405-10	VN085843.D	24 Feb 2025 17:38	JC\MD	Ok
17	Q1415-06	VN085844.D	24 Feb 2025 18:02	JC\MD	Ok
18	VSTDCCC050	VN085845.D	24 Feb 2025 18:26	JC\MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # VN022525

Review By	John Carlone	Review On	2/26/2025 9:38:40 AM
Supervise By	Maresh Dadoda	Supervise On	2/26/2025 1:46:27 PM
SubDirectory	VN022525	HP Acquire Method	HP Processing Method 82N021825W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133142		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133144,VP133145		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VN085846.D	25 Feb 2025 10:11	JC\MD	Ok
2	VSTDCCC050	VN085847.D	25 Feb 2025 10:45	JC\MD	Ok,M
3	VN0225MBL01	VN085848.D	25 Feb 2025 11:19	JC\MD	Ok
4	VN0225WBL01	VN085849.D	25 Feb 2025 11:43	JC\MD	Ok
5	VN0225WBS01	VN085850.D	25 Feb 2025 12:14	JC\MD	Ok,M
6	VN0225WBSD01	VN085851.D	25 Feb 2025 12:48	JC\MD	Ok,M
7	IBLK	VN085852.D	25 Feb 2025 13:12	JC\MD	Ok
8	Q1415-05	VN085853.D	25 Feb 2025 13:36	JC\MD	Ok
9	VN0225MBS01	VN085854.D	25 Feb 2025 14:00	JC\MD	Ok,M
10	Q1395-01	VN085855.D	25 Feb 2025 14:24	JC\MD	Dilution
11	IBLK	VN085856.D	25 Feb 2025 14:49	JC\MD	Ok
12	IBLK	VN085857.D	25 Feb 2025 15:13	JC\MD	Ok
13	Q1395-01DL	VN085858.D	25 Feb 2025 15:37	JC\MD	Ok
14	IBLK	VN085859.D	25 Feb 2025 16:15	JC\MD	Ok
15	VSTDCCC050	VN085860.D	25 Feb 2025 16:42	JC\MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QCBatch ID # VY022525

Review By	Mahesh Dadoda	Review On	2/26/2025 1:45:35 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/26/2025 1:46:55 PM		
SubDirectory	VY022525	HP Acquire Method	MSVOA_Y	HP Processing Method	82y022525s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP133154			
Initial Calibration Stds		VP133137,VP133140,VP133143,VP133146,VP133147,VP133149			
CCC					
Internal Standard/PEM		VP131783			
ICV/I.BLK		VP133150			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY021308.D	25 Feb 2025 12:10	SY/MD	Ok
2	VSTDICC005	VY021309.D	25 Feb 2025 12:40	SY/MD	Ok,M
3	VSTDICC010	VY021310.D	25 Feb 2025 13:03	SY/MD	Ok,M
4	VSTDICC020	VY021311.D	25 Feb 2025 13:26	SY/MD	Ok,M
5	VSTDICCC050	VY021312.D	25 Feb 2025 14:48	SY/MD	Ok,M
6	VSTDICC100	VY021313.D	25 Feb 2025 15:34	SY/MD	Not Ok
7	VSTDICC150	VY021314.D	25 Feb 2025 15:57	SY/MD	Ok,M
8	VIBLK	VY021315.D	25 Feb 2025 16:24	SY/MD	Ok
9	VSTDICC100	VY021316.D	25 Feb 2025 16:49	SY/MD	Ok,M
10	VSTDICV050	VY021317.D	25 Feb 2025 17:31	SY/MD	Ok,M
11	VIBLK	VY021318.D	25 Feb 2025 17:54	SY/MD	Ok

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY022625

Review By	Mahesh Dadoda	Review On	2/27/2025 12:12:40 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/27/2025 12:13:53 PM		
SubDirectory	VY022625	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y022525S.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133161			
CCC		VP133162,VP133163			
Internal Standard/PEM		VP131783			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY021319.D	26 Feb 2025 09:12	SY/MD	Ok
2	VSTDCCC050	VY021320.D	26 Feb 2025 10:11	SY/MD	Ok,M
3	VY0226SBL01	VY021321.D	26 Feb 2025 10:50	SY/MD	Ok
4	VY0226SBS01	VY021322.D	26 Feb 2025 11:22	SY/MD	Ok,M
5	VY0226SBSD01	VY021323.D	26 Feb 2025 11:44	SY/MD	Ok,M
6	Q1416-01	VY021324.D	26 Feb 2025 12:20	SY/MD	ReRun
7	Q1413-02	VY021325.D	26 Feb 2025 12:43	SY/MD	ReRun
8	Q1418-01	VY021326.D	26 Feb 2025 13:07	SY/MD	Ok
9	Q1413-02RE	VY021327.D	26 Feb 2025 13:30	SY/MD	Confirms
10	Q1401-07	VY021328.D	26 Feb 2025 13:54	SY/MD	Ok
11	Q1401-08	VY021329.D	26 Feb 2025 14:17	SY/MD	Ok
12	Q1401-09	VY021330.D	26 Feb 2025 14:40	SY/MD	Ok
13	Q1382-19	VY021331.D	26 Feb 2025 15:04	SY/MD	Ok
14	Q1383-09	VY021332.D	26 Feb 2025 15:27	SY/MD	Ok
15	Q1411-03	VY021333.D	26 Feb 2025 15:51	SY/MD	Ok
16	Q1411-01	VY021334.D	26 Feb 2025 16:14	SY/MD	Not Ok
17	Q1415-02	VY021335.D	26 Feb 2025 16:37	SY/MD	Ok
18	Q1415-01	VY021336.D	26 Feb 2025 17:01	SY/MD	Ok
19	Q1415-04	VY021337.D	26 Feb 2025 17:24	SY/MD	Ok,M
20	Q1415-03	VY021338.D	26 Feb 2025 17:48	SY/MD	Ok
21	Q1420-07	VY021339.D	26 Feb 2025 18:11	SY/MD	ReRun

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY022625

Review By	Mahesh Dadoda	Review On	2/27/2025 12:12:40 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/27/2025 12:13:53 PM		
SubDirectory	VY022625	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y022525S.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133161			
CCC		VP133162,VP133163			
Internal Standard/PEM		VP131783			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	Q1420-03	VY021340.D	26 Feb 2025 18:34	SY/MD	Ok
23	Q1427-01	VY021341.D	26 Feb 2025 18:58	SY/MD	ReRun
24	Q1380-07MS	VY021342.D	26 Feb 2025 19:21	SY/MD	Ok,M
25	Q1380-08MSD	VY021343.D	26 Feb 2025 19:43	SY/MD	Ok,M
26	VSTDCCC050	VY021344.D	26 Feb 2025 20:06	SY/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY022725

Review By	Mahesh Dadoda	Review On	2/28/2025 11:07:46 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:02 AM
SubDirectory	VY022725	HP Acquire Method	HP Processing Method 82y022225s.m
STD. NAME	STD REF.#		
Tune/Reschk	VP133167		
Initial Calibration Stds	VP133165,VP133166		
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY021345.D	27 Feb 2025 09:41	SY/MD	Ok
2	VSTDCCC050	VY021346.D	27 Feb 2025 10:13	SY/MD	Ok,M
3	VY0227SBL01	VY021347.D	27 Feb 2025 10:42	SY/MD	Ok
4	VY0227SBS01	VY021348.D	27 Feb 2025 11:18	SY/MD	Ok,M
5	VY0227SBSD01	VY021349.D	27 Feb 2025 11:40	SY/MD	Ok,M
6	Q1428-01	VY021350.D	27 Feb 2025 12:04	SY/MD	ReRun
7	Q1442-03	VY021351.D	27 Feb 2025 12:28	SY/MD	Ok
8	Q1416-01RE	VY021352.D	27 Feb 2025 12:51	SY/MD	Confirms
9	Q1428-01RE	VY021353.D	27 Feb 2025 13:14	SY/MD	Confirms
10	Q1418-01	VY021354.D	27 Feb 2025 13:38	SY/MD	Not Ok
11	Q1415-03RE	VY021355.D	27 Feb 2025 14:01	SY/MD	Confirms
12	Q1422-02	VY021356.D	27 Feb 2025 14:25	SY/MD	ReRun
13	Q1422-01	VY021357.D	27 Feb 2025 14:48	SY/MD	ReRun
14	Q1421-08	VY021358.D	27 Feb 2025 15:12	SY/MD	Not Ok
15	Q1421-07	VY021359.D	27 Feb 2025 15:35	SY/MD	ReRun
16	Q1421-01	VY021360.D	27 Feb 2025 15:58	SY/MD	Ok
17	Q1427-01RE	VY021361.D	27 Feb 2025 16:22	SY/MD	Confirms
18	Q1420-07	VY021362.D	27 Feb 2025 16:45	SY/MD	Ok
19	Q1411-03	VY021363.D	27 Feb 2025 17:09	SY/MD	Not Ok
20	Q1411-01	VY021364.D	27 Feb 2025 17:32	SY/MD	Ok
21	Q1440-01	VY021365.D	27 Feb 2025 17:55	SY/MD	ReRun

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY022725

Review By	Maresh Dadoda	Review On	2/28/2025 11:07:46 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:02 AM
SubDirectory	VY022725	HP Acquire Method	HP Processing Method 82y022225s.m
STD. NAME	STD REF.#		
Tune/Reschk	VP133167		
Initial Calibration Stds	VP133165,VP133166		
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1448-01	VY021366.D	27 Feb 2025 18:19	SY/MD	ReRun
23	Q1448-02	VY021367.D	27 Feb 2025 18:42	SY/MD	ReRun
24	Q1448-03	VY021368.D	27 Feb 2025 19:06	SY/MD	ReRun
25	Q1448-04	VY021369.D	27 Feb 2025 19:29	SY/MD	Dilution
26	Q1448-05	VY021370.D	27 Feb 2025 19:52	SY/MD	ReRun
27	Q1421-02MS	VY021371.D	27 Feb 2025 20:15	SY/MD	Ok,M
28	Q1421-03MSD	VY021372.D	27 Feb 2025 20:38	SY/MD	Ok,M
29	VSTDCCC050	VY021373.D	27 Feb 2025 21:01	SY/MD	Ok,M

M : Manual Integration