

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN021825

Review By	John Carlone	Review On	2/19/2025 9:44:55 AM
Supervise By	Maresh Dadoda	Supervise On	2/19/2025 1:34:51 PM
SubDirectory	VN021825	HP Acquire Method	HP Processing Method 82N021825W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133068 VP133070,VP133071,VP133072,VP133073,VP133074,VP133075 VP133076

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN085771.D	18 Feb 2025 10:35		JC\MD	Ok
2	VSTDICC100	VSTDICC100	VN085772.D	18 Feb 2025 11:09	Comp.#43 is on Linear Regression	JC\MD	Ok,M
3	VSTDICCC050	VSTDICCC050	VN085773.D	18 Feb 2025 11:32	Comp.#56,58,70,86 is on Quadratic Regression	JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN085774.D	18 Feb 2025 11:56	Not used	JC\MD	Not Ok
5	VSTDICC010	VSTDICC010	VN085775.D	18 Feb 2025 12:20		JC\MD	Ok,M
6	VSTDICC005	VSTDICC005	VN085776.D	18 Feb 2025 12:43	%D failed for Comp. #58 in 01PPB, 05PPB and 20PPB	JC\MD	Ok,M
7	VSTDICC001	VSTDICC001	VN085777.D	18 Feb 2025 13:07		JC\MD	Ok,M
8	IBLK	IBLK	VN085778.D	18 Feb 2025 13:54		JC\MD	Ok
9	VSTDICC020	VSTDICC020	VN085779.D	18 Feb 2025 14:18		JC\MD	Ok,M
10	VSTDICV050	ICVVN021825	VN085780.D	18 Feb 2025 16:28	ICV Failed for comp. #95	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN022425

Review By	John Carlone	Review On	2/25/2025 9:55:02 AM
Supervise By	Maresh Dadoda	Supervise On	2/25/2025 2:29:22 PM
SubDirectory	VN022425	HP Acquire Method	HP Processing Method 82N021825W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133126
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133127,VP133128

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN085828.D	24 Feb 2025 10:50		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN085829.D	24 Feb 2025 11:37	V13516	JC\MD	Ok,M
3	VN0224MBL01	VN0224MBL01	VN085830.D	24 Feb 2025 12:15		JC\MD	Ok
4	VN0224WBL01	VN0224WBL01	VN085831.D	24 Feb 2025 12:39		JC\MD	Ok
5	VN0224WBS01	VN0224WBS01	VN085832.D	24 Feb 2025 13:03		JC\MD	Ok,M
6	VN0224WBSD01	VN0224WBSD01	VN085833.D	24 Feb 2025 13:36		JC\MD	Ok,M
7	Q1405-02DL	MW-04-021325DL	VN085834.D	24 Feb 2025 14:00	vial B pH<2	JC\MD	Ok
8	Q1405-03	MW-20D-021325	VN085835.D	24 Feb 2025 14:24	vial B pH<2	JC\MD	Ok
9	Q1405-04	MW-20D-FD-021325	VN085836.D	24 Feb 2025 14:49	vial B pH<2	JC\MD	Ok,M
10	Q1405-05	MW-22-021925	VN085837.D	24 Feb 2025 15:13	vial B pH<2	JC\MD	Ok
11	Q1405-11	TB-02-021325	VN085838.D	24 Feb 2025 15:37	vial A pH<2 TB	JC\MD	Ok
12	Q1415-05	FB02222025	VN085839.D	24 Feb 2025 16:01	vial A pH<2 FB;hit of com.#16	JC\MD	ReRun
13	Q1405-07	MW-26-021825	VN085840.D	24 Feb 2025 16:26	vial A pH<2	JC\MD	Ok
14	Q1405-08	MOS-11R-021325	VN085841.D	24 Feb 2025 16:50	vial A pH<2	JC\MD	Ok
15	Q1405-09	MOS-13-021925	VN085842.D	24 Feb 2025 17:14	vial A pH<2	JC\MD	Ok
16	Q1405-10	MW-30S-021825	VN085843.D	24 Feb 2025 17:38	vial A pH<2	JC\MD	Ok
17	Q1415-06	B-173-GW01	VN085844.D	24 Feb 2025 18:02	vial A pH<2	JC\MD	Ok
18	VSTDCCC050	VSTDCCC050EC	VN085845.D	24 Feb 2025 18:26		JC\MD	Ok,M

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN022425

Review By	John Carlone	Review On	2/25/2025 9:55:02 AM
Supervise By	Mahesh Dadoda	Supervise On	2/25/2025 2:29:22 PM
SubDirectory	VN022425	HP Acquire Method	HP Processing Method 82N021825W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133126		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133127,VP133128		

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN022525

Review By	John Carlone	Review On	2/26/2025 9:38:40 AM
Supervise By	Maresh Dadoda	Supervise On	2/26/2025 1:46:27 PM
SubDirectory	VN022525	HP Acquire Method	HP Processing Method 82N021825W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133142		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133144,VP133145		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN085846.D	25 Feb 2025 10:11		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN085847.D	25 Feb 2025 10:45	V13516	JC\MD	Ok,M
3	VN0225MBL01	VN0225MBL01	VN085848.D	25 Feb 2025 11:19		JC\MD	Ok
4	VN0225WBL01	VN0225WBL01	VN085849.D	25 Feb 2025 11:43		JC\MD	Ok
5	VN0225WBS01	VN0225WBS01	VN085850.D	25 Feb 2025 12:14		JC\MD	Ok,M
6	VN0225WBSD01	VN0225WBSD01	VN085851.D	25 Feb 2025 12:48		JC\MD	Ok,M
7	IBLK	IBLK	VN085852.D	25 Feb 2025 13:12		JC\MD	Ok
8	Q1415-05	FB02222025	VN085853.D	25 Feb 2025 13:36	vial B pH<2 FB	JC\MD	Ok
9	VN0225MBS01	VN0225MBS01	VN085854.D	25 Feb 2025 14:00		JC\MD	Ok,M
10	Q1395-01	BEL-25-0005	VN085855.D	25 Feb 2025 14:24	Need 10X	JC\MD	Dilution
11	IBLK	IBLK	VN085856.D	25 Feb 2025 14:49		JC\MD	Ok
12	IBLK	IBLK	VN085857.D	25 Feb 2025 15:13		JC\MD	Ok
13	Q1395-01DL	BEL-25-0005DL	VN085858.D	25 Feb 2025 15:37		JC\MD	Ok
14	IBLK	IBLK	VN085859.D	25 Feb 2025 16:15		JC\MD	Ok
15	VSTDCCC050	VSTDCCC050EC	VN085860.D	25 Feb 2025 16:42		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022525

Review By	Maresh Dadoda	Review On	2/26/2025 1:45:35 PM
Supervise By	Semsettin Yesilyurt	Supervise On	2/26/2025 1:46:55 PM
SubDirectory	VY022525	HP Acquire Method	MSVOA_Y HP Processing Method 82y022525s.m

STD. NAME	STD REF.#
Tune/Reschk	VP133154
Initial Calibration Stds	VP133137,VP133140,VP133143,VP133146,VP133147,VP133149
CCC	
Internal Standard/PEM	VP131783
ICV/I.BLK	VP133150
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021308.D	25 Feb 2025 12:10		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY021309.D	25 Feb 2025 12:40		SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY021310.D	25 Feb 2025 13:03		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY021311.D	25 Feb 2025 13:26	Acrolein fail for method	SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021312.D	25 Feb 2025 14:48		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY021313.D	25 Feb 2025 15:34	Not used	SY/MD	Not Ok
7	VSTDIC150	VSTDIC150	VY021314.D	25 Feb 2025 15:57		SY/MD	Ok,M
8	VIBLK	VIBLK	VY021315.D	25 Feb 2025 16:24		SY/MD	Ok
9	VSTDIC100	VSTDIC100	VY021316.D	25 Feb 2025 16:49		SY/MD	Ok,M
10	VSTDICV050	ICVVY022525	VY021317.D	25 Feb 2025 17:31	ICV Failed for comp. #10	SY/MD	Ok,M
11	VIBLK	VIBLK	VY021318.D	25 Feb 2025 17:54		SY/MD	Ok

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022625

Review By	Maresh Dadoda	Review On	2/27/2025 12:12:40 PM
Supervise By	Semsettin Yesilyurt	Supervise On	2/27/2025 12:13:53 PM
SubDirectory	VY022625	HP Acquire Method	MSVOA_Y HP Processing Method 82Y022525S.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133161
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133162,VP133163 VP131783

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021319.D	26 Feb 2025 09:12		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021320.D	26 Feb 2025 10:11		SY/MD	Ok,M
3	VY0226SBL01	VY0226SBL01	VY021321.D	26 Feb 2025 10:50		SY/MD	Ok
4	VY0226SBS01	VY0226SBS01	VY021322.D	26 Feb 2025 11:22		SY/MD	Ok,M
5	VY0226SBSD01	VY0226SBSD01	VY021323.D	26 Feb 2025 11:44		SY/MD	Ok,M
6	Q1416-01	OK-01-022425	VY021324.D	26 Feb 2025 12:20	vial-A ISTD Fail	SY/MD	ReRun
7	Q1413-02	BU-2-022125	VY021325.D	26 Feb 2025 12:43	vial-A ISTD Fail	SY/MD	ReRun
8	Q1418-01	TR-06-02242025	VY021326.D	26 Feb 2025 13:07	vial-A ISTD Fail	SY/MD	Ok
9	Q1413-02RE	BU-2-022125RE	VY021327.D	26 Feb 2025 13:30	vial-B ISTD Fail	SY/MD	Confirms
10	Q1401-07	BP-VPB-192-GW-825-8	VY021328.D	26 Feb 2025 13:54	vial-A	SY/MD	Ok
11	Q1401-08	BP-VPB-192-GW-860-8	VY021329.D	26 Feb 2025 14:17	vial-A	SY/MD	Ok
12	Q1401-09	BP-VPB-192-GW-880-8	VY021330.D	26 Feb 2025 14:40	vial-A	SY/MD	Ok
13	Q1382-19	OU4-PCS-TC-10-02172	VY021331.D	26 Feb 2025 15:04	vial-A	SY/MD	Ok
14	Q1383-09	OU4-PCS-TC-15-02172	VY021332.D	26 Feb 2025 15:27	vial-B	SY/MD	Ok
15	Q1411-03	HR-03-02212025	VY021333.D	26 Feb 2025 15:51	vial-A	SY/MD	Ok
16	Q1411-01	HR-02-02212025	VY021334.D	26 Feb 2025 16:14	vial-A NOT PURGE	SY/MD	Not Ok
17	Q1415-02	B-172-SB01	VY021335.D	26 Feb 2025 16:37	vial-A	SY/MD	Ok
18	Q1415-01	B-163-SB01	VY021336.D	26 Feb 2025 17:01	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022625

Review By	Mahesh Dadoda	Review On	2/27/2025 12:12:40 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/27/2025 12:13:53 PM		
SubDirectory	VY022625	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y022525S.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133161			
CCC		VP133162,VP133163			
Internal Standard/PEM		VP131783			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q1415-04	B-172-SB02	VY021337.D	26 Feb 2025 17:24	vial-A	SY/MD	Ok,M
20	Q1415-03	B-163-SB02	VY021338.D	26 Feb 2025 17:48	vial-A Internal Standard Fail	SY/MD	Ok
21	Q1420-07	TP-2-WC-VOC	VY021339.D	26 Feb 2025 18:11	vial-A Internal Standard Fail	SY/MD	ReRun
22	Q1420-03	TP-1-WC-VOC	VY021340.D	26 Feb 2025 18:34	vial-A	SY/MD	Ok
23	Q1427-01	VNJ-227	VY021341.D	26 Feb 2025 18:58	vial-A Surrogate Fail	SY/MD	ReRun
24	Q1380-07MS	BP-VPB-192-GW-725-7	VY021342.D	26 Feb 2025 19:21	vial-B	SY/MD	Ok,M
25	Q1380-08MSD	BP-VPB-192-GW-725-7	VY021343.D	26 Feb 2025 19:43	vial-B Internal Standard Fail	SY/MD	Ok,M
26	VSTDCCC050	VSTDCCC050EC	VY021344.D	26 Feb 2025 20:06		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022725

Review By	Mahesh Dadoda	Review On	2/28/2025 11:07:46 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:02 AM
SubDirectory	VY022725	HP Acquire Method	HP Processing Method 82y022225s.m

STD. NAME	STD REF.#
Tune/Reschk	VP133167
Initial Calibration Stds	VP133165,VP133166
CCC	
Internal Standard/PEM	
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021345.D	27 Feb 2025 09:41		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021346.D	27 Feb 2025 10:13		SY/MD	Ok,M
3	VY0227SBL01	VY0227SBL01	VY021347.D	27 Feb 2025 10:42		SY/MD	Ok
4	VY0227SBS01	VY0227SBS01	VY021348.D	27 Feb 2025 11:18		SY/MD	Ok,M
5	VY0227SBSD01	VY0227SBSD01	VY021349.D	27 Feb 2025 11:40		SY/MD	Ok,M
6	Q1428-01	NB-07-022525	VY021350.D	27 Feb 2025 12:04	Vial-A ISTD Fail	SY/MD	ReRun
7	Q1442-03	351	VY021351.D	27 Feb 2025 12:28	Vial-A	SY/MD	Ok
8	Q1416-01RE	OK-01-022425RE	VY021352.D	27 Feb 2025 12:51	Vial-B ISTD Fail	SY/MD	Confirms
9	Q1428-01RE	NB-07-022525RE	VY021353.D	27 Feb 2025 13:14	Vial-B Internal Standard Fail	SY/MD	Confirms
10	Q1418-01	TR-06-02242025	VY021354.D	27 Feb 2025 13:38	Vial-B Not purge	SY/MD	Not Ok
11	Q1415-03RE	B-163-SB02RE	VY021355.D	27 Feb 2025 14:01	Vial-B Internal standard fail;Not match for com.#20	SY/MD	Confirms
12	Q1422-02	B-154-SB02	VY021356.D	27 Feb 2025 14:25	Vial-A Internal standard fail	SY/MD	ReRun
13	Q1422-01	B-154-SB01	VY021357.D	27 Feb 2025 14:48	Vial-A Internal standard fail	SY/MD	ReRun
14	Q1421-08	P001-CLAY-CF01-02	VY021358.D	27 Feb 2025 15:12	Vial-B Typo	SY/MD	Not Ok
15	Q1421-07	P001-CLAY-CF01-02	VY021359.D	27 Feb 2025 15:35	Vial-A Internal standard fail	SY/MD	ReRun
16	Q1421-01	P001-CLAY-CF01-01	VY021360.D	27 Feb 2025 15:58	Vial-A Internal standard fail	SY/MD	Ok
17	Q1427-01RE	VNJ-227RE	VY021361.D	27 Feb 2025 16:22	Vial-B Internal standard fail;Surrogate fail	SY/MD	Confirms

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY022725

Review By	Maresh Dadoda	Review On	2/28/2025 11:07:46 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:02 AM
SubDirectory	VY022725	HP Acquire Method	HP Processing Method 82y022225s.m
STD. NAME	STD REF.#		
Tune/Reschk	VP133167		
Initial Calibration Stds	VP133165,VP133166		
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q1420-07	TP-2-WC-VOC	VY021362.D	27 Feb 2025 16:45	Vial-B	SY/MD	Ok
19	Q1411-03	HR-03-02212025	VY021363.D	27 Feb 2025 17:09	Vial-B Not purge	SY/MD	Not Ok
20	Q1411-01	HR-02-02212025	VY021364.D	27 Feb 2025 17:32	Vial-B Internal standard fail	SY/MD	Ok
21	Q1440-01	OR-02-022625	VY021365.D	27 Feb 2025 17:55	Vial-A Internal standard fail	SY/MD	ReRun
22	Q1448-01	PSP1	VY021366.D	27 Feb 2025 18:19	Vial-A Internal standard fail	SY/MD	ReRun
23	Q1448-02	P1	VY021367.D	27 Feb 2025 18:42	Vial-A Internal standard fail	SY/MD	ReRun
24	Q1448-03	P2	VY021368.D	27 Feb 2025 19:06	Vial-A Internal standard fail	SY/MD	ReRun
25	Q1448-04	P3	VY021369.D	27 Feb 2025 19:29	Vial-A Internal standard fail;Need MeOH	SY/MD	Dilution
26	Q1448-05	P4	VY021370.D	27 Feb 2025 19:52	Vial-A Internal standard fail	SY/MD	ReRun
27	Q1421-02MS	P001-CLAY-CF01-01M	VY021371.D	27 Feb 2025 20:15	Vial-A Internal standard fail	SY/MD	Ok,M
28	Q1421-03MSD	P001-CLAY-CF01-01M	VY021372.D	27 Feb 2025 20:38	Vial-A Internal standard fail	SY/MD	Ok,M
29	VSTDCCC050	VSTDCCC050EC	VY021373.D	27 Feb 2025 21:01		SY/MD	Ok,M

M : Manual Integration