

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1415 SAS No.: Q1415 SDG NO.: Q1415
 Lab File ID: VN085829.D Date Analyzed: 02/24/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:37
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	270391	8.22	444288	9.10	401749	11.87
UPPER LIMIT	540782	8.718	888576	9.6	803498	12.365
LOWER LIMIT	135196	7.718	222144	8.6	200875	11.365
EPA SAMPLE NO.						
B-173-GW01	217128	8.22	416577	9.10	372406	11.87
VN0224WBL01	219652	8.22	415867	9.10	368217	11.87
VN0224WBS01	257358	8.22	436252	9.10	378131	11.87
VN0224WBSD01	254490	8.22	420351	9.10	372649	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.