

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1415 SAS No.: Q1415 SDG NO.: Q1415
 Lab File ID: VN085847.D Date Analyzed: 02/25/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:45
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	262718	8.22	419879	9.10	388712	11.86
UPPER LIMIT	525436	8.723	839758	9.6	777424	12.364
LOWER LIMIT	131359	7.723	209940	8.6	194356	11.364
EPA SAMPLE NO.						
FB02222025	210441	8.22	393207	9.10	337792	11.87
VN0225WBL01	216938	8.22	423956	9.10	361345	11.87
VN0225WBS01	265043	8.22	437697	9.10	395653	11.87

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.