

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1415 SAS No.: Q1415 SDG NO.: Q1415
 Lab File ID: VN085847.D Date Analyzed: 02/25/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:45
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	207967	13.788				
UPPER LIMIT	415934	14.288				
LOWER LIMIT	103984	13.288				
EPA SAMPLE NO.						
FB02222025	121445	13.79				
VN0225WBL01	136612	13.79				
VN0225WBS01	197432	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.