

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: Q1415 SAS No.: Q1415 SDG NO.: Q1415
 Lab File ID: VY021346.D Date Analyzed: 02/27/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:13
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	146841	13.347				
UPPER LIMIT	293682	13.847				
LOWER LIMIT	73420.5	12.847				
EPA SAMPLE NO.						
B-163-SB02RE	72948 *	13.35				
VY0227SBL01	107368	13.35				
VY0227SBS01	138215	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.