



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1417

Project: Linden Gen - NJDEP WW Discharge Permit

RunNo.: LB134797

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	02/25/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.505	0.5	101	90-110	02/25/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	02/25/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1417

Project: Linden Gen - NJDEP WW Discharge Permit

RunNo.: LB134857

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Phenolics	ICV1	mg/L	0.97	1	97	90-110	02/28/2025
Sample ID: Phenolics	CCV1	mg/L	0.97	1	97	90-110	02/28/2025
Sample ID: Phenolics	CCV2	mg/L	0.96	1	96	90-110	02/28/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1417

Project: Linden Gen - NJDEP WW Discharge Permit

RunNo.: LB134871

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7.01	7	100	90-110	02/24/2025
Sample ID:	CCV1						
Field pH		pH	7.00	7	100	90-110	02/24/2025
Sample ID:	CCV2						
Field pH		pH	7.01	7	100	90-110	02/24/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1417

Project: Linden Gen - NJDEP WW Discharge Permit

RunNo.: LB134871

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------