

Duplicate Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q1417
Project:	Linden Gen - NJDEP WW Discharge Permit	Sample ID:	LB134845BS
Client ID:	LB134845BSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TPH	mg/L	+/-18	16.8		17.0		1	1.18		02/28/2025

Duplicate Sample Summary

Client: Parkway Generation Operating LLC	SDG No.: Q1417
Project: Linden Gen - NJDEP WW Discharge Permit	Sample ID: Q1417-01
Client ID: LRSA-MOD-1DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.0030	U	0.0030	U	1	0		02/25/2025
Sulfide	mg/L	+/-20	0.43	U	0.43	U	1	0		02/26/2025
Phenolics	mg/L	+/-20	0.015	U	0.015	U	1	0		02/28/2025

Duplicate Sample Summary

Client: Parkway Generation Operating LLC	SDG No.: Q1417
Project: Linden Gen - NJDEP WW Discharge Permit	Sample ID: Q1417-01
Client ID: LRSA-MOD-1MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.99		0.99		2	0.4		02/25/2025
Sulfide	mg/L	+/-20	24.0		23.8		1	0.84		02/26/2025
Phenolics	mg/L	+/-20	1.10		1.00		1	10		02/28/2025

Duplicate Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q1417
Project:	Linden Gen - NJDEP WW Discharge Permit	Sample ID:	Q1417-02
Client ID:	LRSA-MOD-2DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Field pH	pH	+/-20	7.65		7.66		1	0.13		02/24/2025

Duplicate Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q1417
Project:	Linden Gen - NJDEP WW Discharge Permit	Sample ID:	Q1466-01
Client ID:	EFFLUENTMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	66.7		66.5		1	0.3		02/28/2025