

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL121624

Review By	SAM	Review On	12/18/2024 8:48:10 AM
Supervise By		Supervise On	
SubDirectory	VL121624	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL121624AIR.M
STD. NAME	STD REF.#		
Tune/Reschk	AP2556		
Initial Calibration Stds	AP2560,AP2561,AP2562		
CCC	AP2560,MDL,AP2561,AP2562		
Internal Standard/PEM	AP2556		
ICV/I.BLK	AP2558		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL041750.D	16 Dec 2024 12:02	SY/MD	Ok
2	VSTDIC0.5	VL041751.D	16 Dec 2024 12:59	SY/MD	Not Ok
3	VSTDICCC010	VL041752.D	16 Dec 2024 13:30	SY/MD	Ok,M
4	VSTDIC002	VL041753.D	16 Dec 2024 14:03	SY/MD	Ok,M
5	VSTDIC001	VL041754.D	16 Dec 2024 14:34	SY/MD	Ok,M
6	VSTDIC0.5	VL041755.D	16 Dec 2024 15:05	SY/MD	Ok,M
7	VSTDIC0.1	VL041756.D	16 Dec 2024 15:37	SY/MD	Ok,M
8	VSTDIC0.03	VL041757.D	16 Dec 2024 16:08	SY/MD	Ok,M
9	VSTDIC015	VL041758.D	16 Dec 2024 16:41	SY/MD	Ok,M
10	VSTDICV010	VL041759.D	16 Dec 2024 18:59	SY/MD	Ok,M
11	VL1216ABL01	VL041760.D	16 Dec 2024 20:20	SY/MD	Ok
12	VL1216ABL02	VL041761.D	16 Dec 2024 20:51	SY/MD	Ok
13	VL1216ABL03	VL041762.D	16 Dec 2024 21:21	SY/MD	Ok
14	MDL 0.4	VL041763.D	16 Dec 2024 21:52	SY/MD	Ok,M
15	MDL 0.4	VL041764.D	16 Dec 2024 22:23	SY/MD	Ok,M
16	MDL 0.4	VL041765.D	16 Dec 2024 22:54	SY/MD	Ok,M
17	MDL 0.03	VL041766.D	16 Dec 2024 23:24	SY/MD	Ok,M
18	MDL 0.03	VL041767.D	16 Dec 2024 23:55	SY/MD	Ok,M
19	VL1216ABS01	VL041768.D	17 Dec 2024 00:57	SY/MD	Ok,M
20	VSTDCCC010	VL041769.D	17 Dec 2024 07:59	SY/MD	Ok,M

M : Manual Integration