

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q1419 SAS No.: Q1419 SDG No.: Q1419

Instrument ID: MSVOA_L Calibration Date/Time: 02/26/2025 08:38

Lab File ID: VL042050.D Init. Calib. Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:56 12:07

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.364	0.342		-6.04	30
1,1-Dichloroethene	0.376	0.324		-13.83	30
cis-1,2-Dichloroethene	1.353	1.254		-7.32	30
1,1,1-Trichloroethane	1.866	1.748		-6.32	30
Trichloroethene	0.427	0.385		-9.84	30
Tetrachloroethene	0.358	0.339		-5.31	30
1-Bromo-4-Fluorobenzene	0.769	0.787		2.34	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.