

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
 Data File : VL042047.D
 Acq On : 25 Feb 2025 12:07
 Operator : SY/MD
 Sample : VSTDICC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC015

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 02/26/2025
 Supervised By : Mahesh Dadoda 02/26/2025

Quant Time: Feb 25 12:35:10 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Feb 25 10:31:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.800	49	159015	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.978	114	401919	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.904	117	351637	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.393 95 273846 9.441 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 94.400%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.502	85	231978	8.474	ppbv	99
3) Chlorodifluoromethane	1.479	51	324821	10.733	ppbv	100
4) Chloromethane	1.537	50	82097	7.718	ppbv	98
5) Vinyl Chloride	1.586	62	81863	6.436	ppbv	99
6) Bromomethane	1.673	94	39366	7.983	ppbv	97
7) Chloroethane	1.709	64	33329	8.039	ppbv	92
8) Dichlorotetrafluoroethane	1.560	85	188716	7.837	ppbv	98
9) Propene	1.489	41	133086	10.201	ppbv	97
10) Heptane	5.004	43	400418	11.052	ppbv	99
11) Trichlorofluoromethane	1.884	101	230938	8.383	ppbv	95
12) 1,1,2-Trichlorotrifluo...	2.149	101	184306	8.579	ppbv	99
13) Ethanol	1.725	45	4281	2.909	ppbv	100
14) Bromoethene	1.787	108	65496	8.164	ppbv	99
15) Acetone	1.842	43	172538	6.574	ppbv	98
16) 1,3-Butadiene	1.615	39	87490	7.027	ppbv	100
17) tert-Butyl alcohol	2.049	59	239569	7.377	ppbv	98
18) 1,1-Dichloroethene	2.042	96	80535	7.845	ppbv	98
19) Isopropyl Alcohol	1.894	45	102365	6.955	ppbv	99
20) Methylene Chloride	2.068	84	66282	6.709	ppbv	98
21) Allyl Chloride	2.104	41	140726	7.406	ppbv	99
22) trans-1,2-Dichloroethene	2.350	96	94321	8.466	ppbv	95
23) Vinyl Acetate	2.473	43	139727	9.831	ppbv	99
24) 1,1-Dichloroethane	2.418	63	180882	7.997	ppbv	99
25) Ethyl Acetate	2.848	43	767353	10.652	ppbv	99
26) Hexane	2.839	57	306864	10.615	ppbv	99
27) Carbon Disulfide	2.159	76	224609	8.261	ppbv	98
28) Methyl tert-Butyl Ether	2.453	73	168058	9.510	ppbv	98
29) Chloroform	2.861	83	409240	9.774	ppbv	100
30) Cyclohexane	3.874	84	256687	10.946	ppbv	98
31) cis-1,2-Dichloroethene	2.732	61	299324	10.476	ppbv	96
32) 1,1,1-Trichloroethane	3.386	97	417615	8.994	ppbv	98
34) 2-Butanone	2.570	43	457602	13.234	ppbv	98
35) Carbon Tetrachloride	3.780	117	410730	10.524	ppbv	96
36) Benzene	3.674	78	602107	12.367	ppbv	97
37) 1,2-Dichloroethane	3.233	62	335275	12.097	ppbv	99
38) Trichloroethene	4.573	130	228298	11.542	ppbv	98
39) 1,2-Dichloropropane	4.337	63	228418	12.797	ppbv	98
40) 1,4-Dioxane	4.609	88	104519	11.253	ppbv #	96
41) Tetrahydrofuran	3.068	42	271202	13.724	ppbv	100
42) Bromodichloromethane	4.512	83	461272	12.285	ppbv	96
43) Methyl Methacrylate	4.907	69	246468	11.105	ppbv	97
44) 2,2,4-Trimethylpentane	4.667	57	1054880	12.763	ppbv	99
45) t-1,3-Dichloropropene	6.444	75	249612	11.914	ppbv	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022525\
 Data File : VL042047.D
 Acq On : 25 Feb 2025 12:07
 Operator : SY/MD
 Sample : VSTDIC015
 Misc : 400mL/MSVOA_L
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC015

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/26/2025
 Supervised By :Mahesh Dadoda 02/26/2025

Quant Time: Feb 25 12:35:10 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Tue Feb 25 10:31:48 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.635	75	317266	11.935	ppbv	98
47) 1,1,2-Trichloroethane	6.609	97	226021	12.032	ppbv	95
48) Dibromochloromethane	7.412	129	369666	11.750	ppbv	99
49) Bromoform	9.503	173	317322	12.297	ppbv	99
50) 4-Methyl-2-Pentanone	5.787	43	650307	12.237	ppbv	99
51) 2-Hexanone	7.461	43	530165	12.198	ppbv #	100
52) Tetrachloroethene	8.247	164	203233	11.968	ppbv	94
53) Toluene	6.959	91	673822	11.615	ppbv	99
54) 1,2-Dibromoethane	7.677	107	309982	11.985	ppbv	95
56) 1,1,1,2-Tetrachloroethane	8.946	131	276326	10.979	ppbv	99
57) Chlorobenzene	8.943	112	494606	11.957	ppbv	98
58) Ethyl Benzene	9.373	91	922265	11.984	ppbv	100
59) m/p-Xylene	9.571	91	1456036m	24.020	ppbv	
60) o-Xylene	9.982	91	712735	11.626	ppbv	99
61) Styrene	9.888	104	363855	12.446	ppbv	99
62) Isopropylbenzene	10.558	105	1068777	11.612	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.982	83	459839	11.626	ppbv	100
64) n-propylbenzene	11.005	120	282364	12.486	ppbv	99
65) tert-Butylbenzene	11.548	119	924810	11.526	ppbv	99
66) Benzyl Chloride	11.633	91	78975	9.819	ppbv	98
67) sec-Butylbenzene	11.785	105	1315730	11.674	ppbv	99
69) p-Isopropyltoluene	11.940	119	1084253	11.752	ppbv	99
70) n-Butylbenzene	12.296	91	1118373	12.306	ppbv	99
71) 2-Chlorotoluene	10.917	91	818204	11.804	ppbv	100
72) 4-Ethyltoluene	11.141	105	880883	12.264	ppbv	99
73) 1,3,5-Trimethylbenzene	11.218	105	738619	11.921	ppbv	98
74) 1,2,4-Trimethylbenzene	11.552	105	784701	11.801	ppbv	91
75) 1,3-Dichlorobenzene	11.623	146	465436	13.003	ppbv	98
76) 1,4-Dichlorobenzene	11.688	146	467249	13.386	ppbv	99
77) 1,2-Dichlorobenzene	11.963	146	449222	13.129	ppbv	98
78) Hexachloro-1,3-Butadiene	13.895	225	351085	12.403	ppbv	100
79) Naphthalene	13.526	128	764990	11.696	ppbv	99
80) Naphthalene,2-methyl-	14.471	142	324431	10.956	ppbv	99
81) 1,2,4-Trichlorobenzene	13.455	180	354990	13.351	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

