

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1419

Client: GFE LLC

Analytical Method: SWTO-15

Datafile : VL042054.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0226ABS01	Vinyl Chloride	10	9.90	ppbv	99			70	130	
	1,1-Dichloroethene	10	9.30	ppbv	93			70	130	
	cis-1,2-Dichloroethene	10	9.30	ppbv	93			70	130	
	1,1,1-Trichloroethane	10	9.20	ppbv	92			70	130	
	Trichloroethene	10	9.10	ppbv	91			70	130	
	Tetrachloroethene	10	9.40	ppbv	94			70	130	