284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q1419 SAS No.: Q1419 SDG NO.: Q1419

Lab File ID: VL042050.D Date Analyzed: 02/26/2025

Instrument ID: MSVOA_L Time Analyzed: 08:38

GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
SV1	0	0.00				
SV1DL	0	0.00				
IA1	0	0.00				
IA1DUP	0	0.00				
VL0226ABL01	0	0.00				
VL0226ABS01	0	0.00				

IS4 =

AREA UPPER LIMIT = +40% of internal standard area
AREA LOWER LIMIT = -40% of internal standard area
RT UPPER LIMIT = +0.33 minutes of internal standard RT
RT LOWER LIMIT = -0.33 minutes of internal standard RT

 # Column used to flag values outside QC limits with an asterisk.
*** Values outside of QC limits.**