



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Weston Solutions, Inc.

Project Name: RFP 905

Project # N/A

Chemtech Project # Q1421

Test Name: VOC-TCLVOA-10

A. Number of Samples and Date of Receipt:

10 Solid samples were received on 02/24/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Cyanide, EPH, Mercury, Metals ICP-TAL, METALS TAL+CN, PCB, Pesticide-TCL, SPLP BNA, SPLP Cyanide, SPLP Extraction, SPLP ICP Metals, SPLP Mercury, SPLP PCB, SPLP Pesticide, SPLP VOA, SPLP ZHE Ext, SVOC-TCL BNA -20 and VOC-TCLVOA-10. This data package contains results for VOC-TCLVOA-10.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_Y were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868. The analysis of VOC-TCLVOA-10 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements except for P001-CLAY-CF01-01, P001-CLAY-CF01-01MS, P001-CLAY-CF01-01MSD Internal failure confirmed with original sample and P001-CLAY-CF01-02 and P001-CLAY-CF01-02RE, the failure sample in Internal standard was reanalyzed to confirm the failure as per method and reported.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds .

The MSD recoveries met the acceptable requirements .

The RPD for {Q1421-03MSD} with File ID: VY021372.D met criteria except for 1,1,2,2-Tetrachloroethane[40%], 1,2,3-Trichlorobenzene[45%], 1,2,4-Trichlorobenzene[43%], 1,2-Dibromo-3-Chloropropane[34%], Bromodichloromethane[21%], Chloroform[23%], cis-1,2-Dichloroethene[23%], Dibromochloromethane[22%], Isopropylbenzene[26%], Methyl Acetate[31%],

Methylcyclohexane[25%], Methylene Chloride[21%] and trans-1,2-Dichloroethene[21%], due to difference in results of MS and MSD.

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration File ID VY021375.D met the requirements except for Acetone is failing high and associate sample having hit but below CRQL therefore no corrective action taken.

The Tuning criteria met requirements.

E. Calculation:

Low Level Soil Calculation in ug/Kg dry weight basis

$$\frac{(A_x)(I_s)(Df)}{(A_{is})(RRF)(W_s)(D)}$$

Where

Ax = Area for the compound to be measured

Ais = Area for the specific internal standard

Is = Amount of internal standard added in nanograms (ng)

RRF = Relative response factor of the initial calibration curve standard.

Df = Dilution factor

Ws= Weight of sample

D= $\frac{100 - \% \text{moisture}}{100}$

F. Additional Comments:

The soil samples results are based on a dry weight basis.

Trip Blank was not provided with this set of samples.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

G. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____