

Initial and Continuing Calibration Blank Summary

Client: Weston Solutions, Inc.

SDG No.: Q1421

Project: RFP 905

RunNo.: LB134798

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	02/25/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	02/25/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	02/25/2025
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	02/25/2025

Initial and Continuing Calibration Blank Summary

Client: Weston Solutions, Inc.

SDG No.: Q1421

Project: RFP 905

RunNo.: LB134832

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0024	0.005	02/27/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0024	0.005	02/27/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.0024	0.005	02/27/2025
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.0024	0.005	02/27/2025