

Matrix Spike Summary

Client:	Weston Solutions, Inc.	SDG No.:	Q1421
Project:	RFP 905	Sample ID:	Q1421-01
Client ID:	P001-CLAY-CF01-01MS	Percent Solids for Spike Sample:	70.3

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	2.70		0.097	J	2.8	1	93		02/25/2025

Matrix Spike Summary

Client:	Weston Solutions, Inc.	SDG No.:	Q1421
Project:	RFP 905	Sample ID:	Q1421-01
Client ID:	P001-CLAY-CF01-01MSD	Percent Solids for Spike Sample:	70.3

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	2.80		0.097	J	2.8	1	97		02/25/2025

Matrix Spike Summary

Client:	Weston Solutions, Inc.	SDG No.:	Q1421
Project:	RFP 905	Sample ID:	Q1421-04
Client ID:	P001-CLAY-CF01-01MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.036		0.0024	U	0.04	1	90		02/27/2025

Matrix Spike Summary

Client:	Weston Solutions, Inc.	SDG No.:	Q1421
Project:	RFP 905	Sample ID:	Q1421-04
Client ID:	P001-CLAY-CF01-01MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.036		0.0024	U	0.04	1	90		02/27/2025