

### Laboratory Control Sample Summary

**Client:** Weston Solutions, Inc.

**SDG No.:** Q1421

**Project:** RFP 905

**Run No.:** LB134798

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | PB166860BS |            |        |                 |            |                 |                     |               |
| Cyanide   | mg/Kg      | 5          | 4.80   |                 | 96         | 1               | 85-115              | 02/25/2025    |

### Laboratory Control Sample Summary

**Client:** Weston Solutions, Inc.

**SDG No.:** Q1421

**Project:** RFP 905

**Run No.:** LB134832

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | PB166908BS |            |        |                 |            |                 |                     |               |
| Cyanide   | mg/L       | 0.1        | 0.096  |                 | 96         | 1               | 85-115              | 02/27/2025    |