

INSTRUMENT ID: EVI

6029E



Accept=Accepted

$$\begin{array}{r} 610 \\ -622 \\ \hline 10 \end{array}$$

LB134807 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Feb 2025 10:36:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	122	-	0.0000	7471B	PPB	26 Feb 2025 10:52:03	S	Std
0.2 - 1	50.2	1554	-	0.2000	7471B	PPB	26 Feb 2025 10:54:19	S	Std
2.5 - 1	52.5	15643	-	2.5000	7471B	PPB	26 Feb 2025 10:58:12	S	Std
5.0 - 1	55	31038	-	5.0000	7471B	PPB	26 Feb 2025 11:00:29	S	Std
7.5 - 1	57.5	44465	-	7.5000	7471B	PPB	26 Feb 2025 11:02:47	S	Std
10.0 - 1	511	60295	-	10.0000	7471B	PPB	26 Feb 2025 11:07:59	S	Std
ICV19 - 1	ICV19	22136	3.6337	-7471B	PPB	26 Feb 2025 11:10:45	U	SMPL	
ICB19 - 1	ICB19	99	-0.0543	-7471B	PPB	26 Feb 2025 11:15:51	U	SMPL	
CCV25 - 1	CCV25	30370	5.0118	-7471B	PPB	26 Feb 2025 11:18:06	U	SMPL	
CCB25 - 1	CCB25	-159	-0.0975	-7471B	PPB	26 Feb 2025 11:20:21	U	SMPL	
CRA - 1	CRA	1456	0.1728	-7471B	PPB	26 Feb 2025 11:22:39	U	SMPL	
HighStd - 1	HighStd	57631	9.5741	-7471B	PPB	26 Feb 2025 11:24:53	U	SMPL	
ChkStd - 1	ChkStd	37524	6.2090	-7471B	PPB	26 Feb 2025 11:27:09	U	SMPL	
PB166864BL - 1	PBS	175	-0.0416	-7471B	PPB	26 Feb 2025 11:32:18	U	SMPL	
PB166864BS - 1	LCS	22118	3.6307	-7471B	PPB	26 Feb 2025 11:34:34	U	SMPL	
Q1415-01 - 1	B-163-SB01	1229	0.1348	-7471B	PPB	26 Feb 2025 11:36:51	U	SMPL	
Q1415-02 - 1	B-172-SB01	673	0.0418	-7471B	PPB	26 Feb 2025 11:39:12	U	SMPL	
Q1415-03 - 1	B-163-SB02	3150	0.4563	-7471B	PPB	26 Feb 2025 11:41:32	U	SMPL	
Q1415-04 - 1	B-172-SB02	1293	0.1455	-7471B	PPB	26 Feb 2025 11:43:50	U	SMPL	
Q1416-01 - 1	OK-01-022425	207	-0.0362	-7471B	PPB	26 Feb 2025 11:46:06	U	SMPL	
CCV26 - 1	CCV26	29434	4.8551	-7471B	PPB	26 Feb 2025 11:48:22	U	SMPL	
CCB26 - 1	CCB26	327	-0.0161	-7471B	PPB	26 Feb 2025 11:53:21	U	SMPL	
Q1418-01 - 1	TR-06-02242025	2013	0.2660	-7471B	PPB	26 Feb 2025 11:55:37	U	SMPL	
Q1420-01 - 1	TP-1-WC	3631	0.5368	-7471B	PPB	26 Feb 2025 11:57:52	U	SMPL	
Q1420-05 - 1	TP-2-WC	1837	0.2366	-7471B	PPB	26 Feb 2025 12:00:08	U	SMPL	
Q1421-01 - 1	P001-CLAY-CF01-01	2242	0.3043	-7471B	PPB	26 Feb 2025 12:02:25	U	SMPL	
Q1421-01DUP - 1	P001-CLAY-CF01-01DUP	2144	0.2879	-7471B	PPB	26 Feb 2025 12:07:56	U	SMPL	
Q1421-02 - 1	Q1421-01MS	31573	5.2131	-7471B	PPB	26 Feb 2025 12:16:29	U	SMPL	
Q1421-03 - 1	Q1421-01MSD	30173	4.9788	-7471B	PPB	26 Feb 2025 12:21:24	U	SMPL	
Q1421-07 - 1	P001-CLAY-CF01-02	491	0.0113	-7471B	PPB	26 Feb 2025 12:23:42	U	SMPL	
Q1421-09 - 1	P001-CLAY-CF02-01	769	0.0578	-7471B	PPB	26 Feb 2025 12:26:02	U	SMPL	
Q1422-01 - 1	B-154-SB01	2191	0.2958	-7471B	PPB	26 Feb 2025 12:28:18	U	SMPL	
CCV27 - 1	CCV27	28956	4.7751	-7471B	PPB	26 Feb 2025 12:30:34	U	SMPL	
CCB27 - 1	CCB27	147	-0.0463	-7471B	PPB	26 Feb 2025 12:37:47	U	SMPL	
Q1422-02 - 1	B-154-SB02	708	0.0476	-7471B	PPB	26 Feb 2025 12:40:02	U	SMPL	
Q1421-01LX5 - 1		239	-0.0309	-7471B	PPB	26 Feb 2025 12:42:18	U	SMPL	
Q1421-01A - 1		28517	4.7017	-7471B	PPB	26 Feb 2025 12:44:34	U	SMPL	
CCV28 - 1	CCV28	28686	4.7299	-7471B	PPB	26 Feb 2025 12:46:50	U	SMPL	
CCB28 - 1	CCB28	-33	-0.0764	-7471B	PPB	26 Feb 2025 12:49:09	U	SMPL	