



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Weston Solutions, Inc. **SDG No.:** Q1421
Contract: ROYF02 **Lab Code:** CHEM **Case No.:** Q1421 **SAS No.:** Q1421
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.17	0.2	86	40 - 160	CV	02/26/2025	11:22	LB134807
CRI01	Aluminum	98.3	100	98	40 - 160	P	03/10/2025	12:01	LB134965
	Antimony	49.7	50.0	100	40 - 160	P	03/10/2025	12:01	LB134965
	Arsenic	18.2	20.0	91	40 - 160	P	03/10/2025	12:01	LB134965
	Barium	90.9	100	91	40 - 160	P	03/10/2025	12:01	LB134965
	Beryllium	5.47	6.0	91	40 - 160	P	03/10/2025	12:01	LB134965
	Cadmium	5.67	6.0	94	40 - 160	P	03/10/2025	12:01	LB134965
	Calcium	1900	2000	95	40 - 160	P	03/10/2025	12:01	LB134965
	Chromium	10.4	10.0	104	40 - 160	P	03/10/2025	12:01	LB134965
	Cobalt	28.3	30.0	94	40 - 160	P	03/10/2025	12:01	LB134965
	Copper	21.4	20.0	107	40 - 160	P	03/10/2025	12:01	LB134965
	Iron	111	100	111	40 - 160	P	03/10/2025	12:01	LB134965
	Lead	12.0	12.0	100	40 - 160	P	03/10/2025	12:01	LB134965
	Magnesium	1980	2000	99	40 - 160	P	03/10/2025	12:01	LB134965
	Manganese	19.3	20.0	96	40 - 160	P	03/10/2025	12:01	LB134965
	Nickel	37.9	40.0	95	40 - 160	P	03/10/2025	12:01	LB134965
	Potassium	1950	2000	97	40 - 160	P	03/10/2025	12:01	LB134965
	Selenium	21.0	20.0	105	40 - 160	P	03/10/2025	12:01	LB134965
	Silver	10.2	10.0	102	40 - 160	P	03/10/2025	12:01	LB134965
	Sodium	1810	2000	91	40 - 160	P	03/10/2025	12:01	LB134965
	Thallium	39.5	40.0	99	40 - 160	P	03/10/2025	12:01	LB134965
	Vanadium	36.7	40.0	92	40 - 160	P	03/10/2025	12:01	LB134965
	Zinc	41.3	40.0	103	40 - 160	P	03/10/2025	12:01	LB134965