



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Weston Solutions, Inc.				SDG No.:	Q1421				
Contract:	ROYF02	Lab Code:	CHEM		Case No.:	Q1421	SAS No.:	Q1421		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB19	Mercury	0.20	+/-0.20	U	0.20	CV	02/26/2025	11:15	LB134807	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q1421

Contract: ROYF02

Lab Code: CHEM

Case No.: Q1421

SAS No.: Q1421

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB25	Mercury	0.20	+/-0.20	U	0.20	CV	02/26/2025	11:20	LB134807
CCB26	Mercury	0.20	+/-0.20	U	0.20	CV	02/26/2025	11:53	LB134807
CCB27	Mercury	0.20	+/-0.20	U	0.20	CV	02/26/2025	12:37	LB134807
CCB28	Mercury	0.20	+/-0.20	U	0.20	CV	02/26/2025	12:49	LB134807

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions, Inc.</u>		SDG No.: <u>Q1421</u>							
Contract: <u>ROYF02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1421</u>	SAS No.: <u>Q1421</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	03/10/2025	11:57	LB134965
	Antimony	50.0	+/-50.0	U	50.0	P	03/10/2025	11:57	LB134965
	Arsenic	20.0	+/-20.0	U	20.0	P	03/10/2025	11:57	LB134965
	Barium	100	+/-100	U	100	P	03/10/2025	11:57	LB134965
	Beryllium	6.00	+/-6.00	U	6.00	P	03/10/2025	11:57	LB134965
	Cadmium	6.00	+/-6.00	U	6.00	P	03/10/2025	11:57	LB134965
	Calcium	2000	+/-2000	U	2000	P	03/10/2025	11:57	LB134965
	Chromium	10.0	+/-10.0	U	10.0	P	03/10/2025	11:57	LB134965
	Cobalt	30.0	+/-30.0	U	30.0	P	03/10/2025	11:57	LB134965
	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	11:57	LB134965
	Iron	100	+/-100	U	100	P	03/10/2025	11:57	LB134965
	Lead	12.0	+/-12.0	U	12.0	P	03/10/2025	11:57	LB134965
	Magnesium	2000	+/-2000	U	2000	P	03/10/2025	11:57	LB134965
	Manganese	20.0	+/-20.0	U	20.0	P	03/10/2025	11:57	LB134965
	Nickel	40.0	+/-40.0	U	40.0	P	03/10/2025	11:57	LB134965
	Potassium	2000	+/-2000	U	2000	P	03/10/2025	11:57	LB134965
	Selenium	20.0	+/-20.0	U	20.0	P	03/10/2025	11:57	LB134965
	Silver	10.0	+/-10.0	U	10.0	P	03/10/2025	11:57	LB134965
	Sodium	2000	+/-2000	U	2000	P	03/10/2025	11:57	LB134965
	Thallium	40.0	+/-40.0	U	40.0	P	03/10/2025	11:57	LB134965
	Vanadium	40.0	+/-40.0	U	40.0	P	03/10/2025	11:57	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	11:57	LB134965

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions, Inc.</u>		SDG No.: <u>Q1421</u>							
Contract: <u>ROYF02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1421</u>	SAS No.: <u>Q1421</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	03/10/2025	12:34	LB134965
	Antimony	50.0	+/-50.0	U	50.0	P	03/10/2025	12:34	LB134965
	Arsenic	20.0	+/-20.0	U	20.0	P	03/10/2025	12:34	LB134965
	Barium	100	+/-100	U	100	P	03/10/2025	12:34	LB134965
	Beryllium	6.00	+/-6.00	U	6.00	P	03/10/2025	12:34	LB134965
	Cadmium	6.00	+/-6.00	U	6.00	P	03/10/2025	12:34	LB134965
	Calcium	2000	+/-2000	U	2000	P	03/10/2025	12:34	LB134965
	Chromium	10.0	+/-10.0	U	10.0	P	03/10/2025	12:34	LB134965
	Cobalt	30.0	+/-30.0	U	30.0	P	03/10/2025	12:34	LB134965
	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	12:34	LB134965
	Iron	100	+/-100	U	100	P	03/10/2025	12:34	LB134965
	Lead	12.0	+/-12.0	U	12.0	P	03/10/2025	12:34	LB134965
	Magnesium	2000	+/-2000	U	2000	P	03/10/2025	12:34	LB134965
	Manganese	20.0	+/-20.0	U	20.0	P	03/10/2025	12:34	LB134965
	Nickel	40.0	+/-40.0	U	40.0	P	03/10/2025	12:34	LB134965
	Potassium	2000	+/-2000	U	2000	P	03/10/2025	12:34	LB134965
	Selenium	20.0	+/-20.0	U	20.0	P	03/10/2025	12:34	LB134965
	Silver	10.0	+/-10.0	U	10.0	P	03/10/2025	12:34	LB134965
	Sodium	2000	+/-2000	U	2000	P	03/10/2025	12:34	LB134965
	Thallium	40.0	+/-40.0	U	40.0	P	03/10/2025	12:34	LB134965
	Vanadium	40.0	+/-40.0	U	40.0	P	03/10/2025	12:34	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	12:34	LB134965
CCB02	Aluminum	100	+/-100	U	100	P	03/10/2025	13:25	LB134965
	Antimony	50.0	+/-50.0	U	50.0	P	03/10/2025	13:25	LB134965
	Arsenic	20.0	+/-20.0	U	20.0	P	03/10/2025	13:25	LB134965
	Barium	100	+/-100	U	100	P	03/10/2025	13:25	LB134965
	Beryllium	6.00	+/-6.00	U	6.00	P	03/10/2025	13:25	LB134965
	Cadmium	6.00	+/-6.00	U	6.00	P	03/10/2025	13:25	LB134965
	Calcium	2000	+/-2000	U	2000	P	03/10/2025	13:25	LB134965
	Chromium	10.0	+/-10.0	U	10.0	P	03/10/2025	13:25	LB134965
	Cobalt	30.0	+/-30.0	U	30.0	P	03/10/2025	13:25	LB134965
	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	13:25	LB134965
	Iron	100	+/-100	U	100	P	03/10/2025	13:25	LB134965
	Lead	12.0	+/-12.0	U	12.0	P	03/10/2025	13:25	LB134965
	Magnesium	2000	+/-2000	U	2000	P	03/10/2025	13:25	LB134965
	Manganese	20.0	+/-20.0	U	20.0	P	03/10/2025	13:25	LB134965
	Nickel	40.0	+/-40.0	U	40.0	P	03/10/2025	13:25	LB134965
	Potassium	2000	+/-2000	U	2000	P	03/10/2025	13:25	LB134965
	Selenium	20.0	+/-20.0	U	20.0	P	03/10/2025	13:25	LB134965

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions, Inc.</u>		SDG No.: <u>Q1421</u>							
Contract: <u>ROYF02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1421</u>	SAS No.: <u>Q1421</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	03/10/2025	13:25	LB134965
	Sodium	2000	+/-2000	U	2000	P	03/10/2025	13:25	LB134965
	Thallium	40.0	+/-40.0	U	40.0	P	03/10/2025	13:25	LB134965
	Vanadium	40.0	+/-40.0	U	40.0	P	03/10/2025	13:25	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	13:25	LB134965
CCB03	Aluminum	100	+/-100	U	100	P	03/10/2025	14:15	LB134965
	Antimony	50.0	+/-50.0	U	50.0	P	03/10/2025	14:15	LB134965
	Arsenic	20.0	+/-20.0	U	20.0	P	03/10/2025	14:15	LB134965
	Barium	100	+/-100	U	100	P	03/10/2025	14:15	LB134965
	Beryllium	6.00	+/-6.00	U	6.00	P	03/10/2025	14:15	LB134965
	Cadmium	6.00	+/-6.00	U	6.00	P	03/10/2025	14:15	LB134965
	Calcium	2000	+/-2000	U	2000	P	03/10/2025	14:15	LB134965
	Chromium	10.0	+/-10.0	U	10.0	P	03/10/2025	14:15	LB134965
	Cobalt	30.0	+/-30.0	U	30.0	P	03/10/2025	14:15	LB134965
	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	14:15	LB134965
	Iron	100	+/-100	U	100	P	03/10/2025	14:15	LB134965
	Lead	12.0	+/-12.0	U	12.0	P	03/10/2025	14:15	LB134965
	Magnesium	2000	+/-2000	U	2000	P	03/10/2025	14:15	LB134965
	Manganese	20.0	+/-20.0	U	20.0	P	03/10/2025	14:15	LB134965
	Nickel	40.0	+/-40.0	U	40.0	P	03/10/2025	14:15	LB134965
	Potassium	2000	+/-2000	U	2000	P	03/10/2025	14:15	LB134965
	Selenium	20.0	+/-20.0	U	20.0	P	03/10/2025	14:15	LB134965
	Silver	10.0	+/-10.0	U	10.0	P	03/10/2025	14:15	LB134965
	Sodium	2000	+/-2000	U	2000	P	03/10/2025	14:15	LB134965
	Thallium	40.0	+/-40.0	U	40.0	P	03/10/2025	14:15	LB134965
	Vanadium	40.0	+/-40.0	U	40.0	P	03/10/2025	14:15	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	14:15	LB134965
CCB04	Aluminum	100	+/-100	U	100	P	03/10/2025	15:05	LB134965
	Antimony	50.0	+/-50.0	U	50.0	P	03/10/2025	15:05	LB134965
	Arsenic	20.0	+/-20.0	U	20.0	P	03/10/2025	15:05	LB134965
	Barium	100	+/-100	U	100	P	03/10/2025	15:05	LB134965
	Beryllium	6.00	+/-6.00	U	6.00	P	03/10/2025	15:05	LB134965
	Cadmium	6.00	+/-6.00	U	6.00	P	03/10/2025	15:05	LB134965
	Calcium	2000	+/-2000	U	2000	P	03/10/2025	15:05	LB134965
	Chromium	10.0	+/-10.0	U	10.0	P	03/10/2025	15:05	LB134965
	Cobalt	30.0	+/-30.0	U	30.0	P	03/10/2025	15:05	LB134965
	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	15:05	LB134965
	Iron	100	+/-100	U	100	P	03/10/2025	15:05	LB134965
	Lead	12.0	+/-12.0	U	12.0	P	03/10/2025	15:05	LB134965

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions, Inc.</u>		SDG No.: <u>Q1421</u>							
Contract: <u>ROYF02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1421</u>	SAS No.: <u>Q1421</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	03/10/2025	15:05	LB134965
	Manganese	20.0	+/-20.0	U	20.0	P	03/10/2025	15:05	LB134965
	Nickel	40.0	+/-40.0	U	40.0	P	03/10/2025	15:05	LB134965
	Potassium	2000	+/-2000	U	2000	P	03/10/2025	15:05	LB134965
	Selenium	20.0	+/-20.0	U	20.0	P	03/10/2025	15:05	LB134965
	Silver	10.0	+/-10.0	U	10.0	P	03/10/2025	15:05	LB134965
	Sodium	2000	+/-2000	U	2000	P	03/10/2025	15:05	LB134965
	Thallium	40.0	+/-40.0	U	40.0	P	03/10/2025	15:05	LB134965
	Vanadium	40.0	+/-40.0	U	40.0	P	03/10/2025	15:05	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	15:05	LB134965
CCB05	Aluminum	100	+/-100	U	100	P	03/10/2025	15:55	LB134965
	Antimony	50.0	+/-50.0	U	50.0	P	03/10/2025	15:55	LB134965
	Arsenic	20.0	+/-20.0	U	20.0	P	03/10/2025	15:55	LB134965
	Barium	100	+/-100	U	100	P	03/10/2025	15:55	LB134965
	Beryllium	6.00	+/-6.00	U	6.00	P	03/10/2025	15:55	LB134965
	Cadmium	6.00	+/-6.00	U	6.00	P	03/10/2025	15:55	LB134965
	Calcium	2000	+/-2000	U	2000	P	03/10/2025	15:55	LB134965
	Chromium	10.0	+/-10.0	U	10.0	P	03/10/2025	15:55	LB134965
	Cobalt	30.0	+/-30.0	U	30.0	P	03/10/2025	15:55	LB134965
	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	15:55	LB134965
	Iron	100	+/-100	U	100	P	03/10/2025	15:55	LB134965
	Lead	12.0	+/-12.0	U	12.0	P	03/10/2025	15:55	LB134965
	Magnesium	2000	+/-2000	U	2000	P	03/10/2025	15:55	LB134965
	Manganese	20.0	+/-20.0	U	20.0	P	03/10/2025	15:55	LB134965
	Nickel	40.0	+/-40.0	U	40.0	P	03/10/2025	15:55	LB134965
	Potassium	2000	+/-2000	U	2000	P	03/10/2025	15:55	LB134965
	Selenium	20.0	+/-20.0	U	20.0	P	03/10/2025	15:55	LB134965
	Silver	10.0	+/-10.0	U	10.0	P	03/10/2025	15:55	LB134965
	Sodium	2000	+/-2000	U	2000	P	03/10/2025	15:55	LB134965
	Thallium	40.0	+/-40.0	U	40.0	P	03/10/2025	15:55	LB134965
	Vanadium	40.0	+/-40.0	U	40.0	P	03/10/2025	15:55	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	15:55	LB134965
CCB06	Aluminum	100	+/-100	U	100	P	03/10/2025	16:52	LB134965
	Antimony	50.0	+/-50.0	U	50.0	P	03/10/2025	16:52	LB134965
	Arsenic	20.0	+/-20.0	U	20.0	P	03/10/2025	16:52	LB134965
	Barium	100	+/-100	U	100	P	03/10/2025	16:52	LB134965
	Beryllium	6.00	+/-6.00	U	6.00	P	03/10/2025	16:52	LB134965
	Cadmium	6.00	+/-6.00	U	6.00	P	03/10/2025	16:52	LB134965
	Calcium	2000	+/-2000	U	2000	P	03/10/2025	16:52	LB134965

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions, Inc.</u>		SDG No.: <u>Q1421</u>							
Contract: <u>ROYF02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1421</u>	SAS No.: <u>Q1421</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	03/10/2025	16:52	LB134965
	Cobalt	30.0	+/-30.0	U	30.0	P	03/10/2025	16:52	LB134965
	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	16:52	LB134965
	Iron	100	+/-100	U	100	P	03/10/2025	16:52	LB134965
	Lead	12.0	+/-12.0	U	12.0	P	03/10/2025	16:52	LB134965
	Magnesium	2000	+/-2000	U	2000	P	03/10/2025	16:52	LB134965
	Manganese	20.0	+/-20.0	U	20.0	P	03/10/2025	16:52	LB134965
	Nickel	40.0	+/-40.0	U	40.0	P	03/10/2025	16:52	LB134965
	Potassium	2000	+/-2000	U	2000	P	03/10/2025	16:52	LB134965
	Selenium	20.0	+/-20.0	U	20.0	P	03/10/2025	16:52	LB134965
	Silver	10.0	+/-10.0	U	10.0	P	03/10/2025	16:52	LB134965
	Sodium	2000	+/-2000	U	2000	P	03/10/2025	16:52	LB134965
	Thallium	40.0	+/-40.0	U	40.0	P	03/10/2025	16:52	LB134965
	Vanadium	40.0	+/-40.0	U	40.0	P	03/10/2025	16:52	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	16:52	LB134965
CCB07	Aluminum	100	+/-100	U	100	P	03/10/2025	17:44	LB134965
	Antimony	50.0	+/-50.0	U	50.0	P	03/10/2025	17:44	LB134965
	Arsenic	20.0	+/-20.0	U	20.0	P	03/10/2025	17:44	LB134965
	Barium	100	+/-100	U	100	P	03/10/2025	17:44	LB134965
	Beryllium	6.00	+/-6.00	U	6.00	P	03/10/2025	17:44	LB134965
	Cadmium	6.00	+/-6.00	U	6.00	P	03/10/2025	17:44	LB134965
	Calcium	2000	+/-2000	U	2000	P	03/10/2025	17:44	LB134965
	Chromium	10.0	+/-10.0	U	10.0	P	03/10/2025	17:44	LB134965
	Cobalt	30.0	+/-30.0	U	30.0	P	03/10/2025	17:44	LB134965
	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	17:44	LB134965
	Iron	100	+/-100	U	100	P	03/10/2025	17:44	LB134965
	Lead	12.0	+/-12.0	U	12.0	P	03/10/2025	17:44	LB134965
	Magnesium	2000	+/-2000	U	2000	P	03/10/2025	17:44	LB134965
	Manganese	20.0	+/-20.0	U	20.0	P	03/10/2025	17:44	LB134965
	Nickel	40.0	+/-40.0	U	40.0	P	03/10/2025	17:44	LB134965
	Potassium	2000	+/-2000	U	2000	P	03/10/2025	17:44	LB134965
	Selenium	20.0	+/-20.0	U	20.0	P	03/10/2025	17:44	LB134965
	Silver	10.0	+/-10.0	U	10.0	P	03/10/2025	17:44	LB134965
	Sodium	2000	+/-2000	U	2000	P	03/10/2025	17:44	LB134965
	Thallium	40.0	+/-40.0	U	40.0	P	03/10/2025	17:44	LB134965
	Vanadium	40.0	+/-40.0	U	40.0	P	03/10/2025	17:44	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	17:44	LB134965
CCB08	Aluminum	100	+/-100	U	100	P	03/10/2025	18:33	LB134965
	Antimony	50.0	+/-50.0	U	50.0	P	03/10/2025	18:33	LB134965

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q1421

Contract: ROYF02

Lab Code: CHEM

Case No.: Q1421

SAS No.: Q1421

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	03/10/2025	18:33	LB134965
	Barium	100	+/-100	U	100	P	03/10/2025	18:33	LB134965
	Beryllium	6.00	+/-6.00	U	6.00	P	03/10/2025	18:33	LB134965
	Cadmium	6.00	+/-6.00	U	6.00	P	03/10/2025	18:33	LB134965
	Calcium	2000	+/-2000	U	2000	P	03/10/2025	18:33	LB134965
	Chromium	10.0	+/-10.0	U	10.0	P	03/10/2025	18:33	LB134965
	Cobalt	30.0	+/-30.0	U	30.0	P	03/10/2025	18:33	LB134965
	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	18:33	LB134965
	Iron	100	+/-100	U	100	P	03/10/2025	18:33	LB134965
	Lead	12.0	+/-12.0	U	12.0	P	03/10/2025	18:33	LB134965
	Magnesium	2000	+/-2000	U	2000	P	03/10/2025	18:33	LB134965
	Manganese	20.0	+/-20.0	U	20.0	P	03/10/2025	18:33	LB134965
	Nickel	40.0	+/-40.0	U	40.0	P	03/10/2025	18:33	LB134965
	Potassium	2000	+/-2000	U	2000	P	03/10/2025	18:33	LB134965
	Selenium	20.0	+/-20.0	U	20.0	P	03/10/2025	18:33	LB134965
	Silver	10.0	+/-10.0	U	10.0	P	03/10/2025	18:33	LB134965
	Sodium	2000	+/-2000	U	2000	P	03/10/2025	18:33	LB134965
	Thallium	40.0	+/-40.0	U	40.0	P	03/10/2025	18:33	LB134965
	Vanadium	40.0	+/-40.0	U	40.0	P	03/10/2025	18:33	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	18:33	LB134965
CCB09	Aluminum	100	+/-100	U	100	P	03/10/2025	18:50	LB134965
	Antimony	50.0	+/-50.0	U	50.0	P	03/10/2025	18:50	LB134965
	Arsenic	20.0	+/-20.0	U	20.0	P	03/10/2025	18:50	LB134965
	Barium	100	+/-100	U	100	P	03/10/2025	18:50	LB134965
	Beryllium	6.00	+/-6.00	U	6.00	P	03/10/2025	18:50	LB134965
	Cadmium	6.00	+/-6.00	U	6.00	P	03/10/2025	18:50	LB134965
	Calcium	2000	+/-2000	U	2000	P	03/10/2025	18:50	LB134965
	Chromium	10.0	+/-10.0	U	10.0	P	03/10/2025	18:50	LB134965
	Cobalt	30.0	+/-30.0	U	30.0	P	03/10/2025	18:50	LB134965
	Copper	20.0	+/-20.0	U	20.0	P	03/10/2025	18:50	LB134965
	Iron	100	+/-100	U	100	P	03/10/2025	18:50	LB134965
	Lead	12.0	+/-12.0	U	12.0	P	03/10/2025	18:50	LB134965
	Magnesium	2000	+/-2000	U	2000	P	03/10/2025	18:50	LB134965
	Manganese	20.0	+/-20.0	U	20.0	P	03/10/2025	18:50	LB134965
	Nickel	40.0	+/-40.0	U	40.0	P	03/10/2025	18:50	LB134965
	Potassium	2000	+/-2000	U	2000	P	03/10/2025	18:50	LB134965
	Selenium	20.0	+/-20.0	U	20.0	P	03/10/2025	18:50	LB134965
	Silver	10.0	+/-10.0	U	10.0	P	03/10/2025	18:50	LB134965
	Sodium	2000	+/-2000	U	2000	P	03/10/2025	18:50	LB134965

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions, Inc.</u>		SDG No.: <u>Q1421</u>							
Contract: <u>ROYF02</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q1421</u>		SAS No.: <u>Q1421</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	40.0	+/-40.0	U	40.0	P	03/10/2025	18:50	LB134965
	Vanadium	40.0	+/-40.0	U	40.0	P	03/10/2025	18:50	LB134965
	Zinc	40.0	+/-40.0	U	40.0	P	03/10/2025	18:50	LB134965