


LB134835

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%SD
0.0	0.000	-0.041	-0.041	60	0.000	60					
0.2	0.200	0.175	-0.025	1020	0.0 %	1020	0				-13
2.5	2.500	2.477	-0.023	11285	0.0 %	11285					-1
5.0	5.000	5.246	0.246	23631	0.0 %	23631					-5
7.5	7.500	7.370	-0.130	33104	0.0 %	33104					-2
10.0	10.000	9.972	-0.028	44705	0.0 %	44705					0

LB134835 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Feb 2025 13:10:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	60	-	0.0000	7470A	PPB	27 Feb 2025 13:12:36	S	Std
0.2 - 1	50-2	1020	-	0.2000	7470A	PPB	27 Feb 2025 13:14:53	S	Std
2.5 - 1	52.5	11285	-	2.5000	7470A	PPB	27 Feb 2025 13:17:10	S	Std
5.0 - 1	55	23631	-	5.0000	7470A	PPB	27 Feb 2025 13:19:27	S	Std
7.5 - 1	57.5	33104	-	7.5000	7470A	PPB	27 Feb 2025 13:21:45	S	Std
10.0 - 1	518	44705	-	10.0000	7470A	PPB	27 Feb 2025 13:24:04	S	Std
ICV21 - 1	ICV21	16824	3.7193	-	7470A	PPB	27 Feb 2025 13:26:59	U	SMPL
ICB21 - 1	ICB21	-200	-0.0988	-	7470A	PPB	27 Feb 2025 13:29:14	U	SMPL
CCV32 - 1	CCV32	23868	5.2991	-	7470A	PPB	27 Feb 2025 13:31:31	U	SMPL
CCB32 - 1	CCB32	56	-0.0414	-	7470A	PPB	27 Feb 2025 13:36:21	U	SMPL
CRA - 1	CRA	1010	0.1725	-	7470A	PPB	27 Feb 2025 13:38:36	U	SMPL
HighStd - 1	HighStd	43869	9.7848	-	7470A	PPB	27 Feb 2025 13:40:51	U	SMPL
ChkStd - 1	ChkStd	28950	6.4389	-	7470A	PPB	27 Feb 2025 13:43:07	U	SMPL
PB166884BL - 1	PBW	72	-0.0378	-	7470A	PPB	27 Feb 2025 13:49:03	U	SMPL
PB166884BS - 1	LCSW	16634	3.6766	-	7470A	PPB	27 Feb 2025 13:51:22	U	SMPL
Q1421-04 - 1	P001-CLAY-CF01-01	343	0.0229	-	7470A	PPB	27 Feb 2025 13:59:10	U	SMPL
Q1421-04DUP - 1	P001-CLAY-CF01-01DUP	209	-0.0071	-	7470A	PPB	27 Feb 2025 14:01:26	U	SMPL
Q1421-05 - 1	Q1421-04MS	0	-0.0540	-	7470A	PPB	27 Feb 2025 14:03:43	U	SMPL
Q1421-05 - 1	Q1421-04MS	16671	3.6849	-	7470A	PPB	27 Feb 2025 14:06:09	U	SMPL
Q1421-06 - 1	Q1421-04MSD	16809	3.7159	-	7470A	PPB	27 Feb 2025 14:08:28	U	SMPL
Q1421-08 - 1	P001-CLAY-CF01-02	57	-0.0412	-	7470A	PPB	27 Feb 2025 14:10:45	U	SMPL
CCV33 - 1	CCV33	23076	5.1214	-	7470A	PPB	27 Feb 2025 14:13:02	U	SMPL
CCB33 - 1	CCB33	-106	-0.0778	-	7470A	PPB	27 Feb 2025 14:15:17	U	SMPL
Q1421-10 - 1	P001-CLAY-CF02-01	240	-0.0002	-	7470A	PPB	27 Feb 2025 14:17:35	U	SMPL
PB166861TB - 1	PB166861TB	141	-0.0224	-	7470A	PPB	27 Feb 2025 14:19:51	U	SMPL
Q1421-04LX5 - 1		59	-0.0408	-	7470A	PPB	27 Feb 2025 14:22:06	U	SMPL
Q1421-04A - 1		16791	3.7119	-	7470A	PPB	27 Feb 2025 14:24:22	U	SMPL
CCV34 - 1	CCV34	23256	5.1618	-	7470A	PPB	27 Feb 2025 14:45:47	U	SMPL
CCB34 - 1	CCB34	486	0.0550	-	7470A	PPB	27 Feb 2025 14:51:10	U	SMPL