

Continuing Calibration Report for SequenceID : FC022725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068305.D

Aliphatic C28-C40	12483547.000	120.000	16.763	21.566	104029.558	102846.462	-1.150
Aliphatic EPH	44248852.000	360.000	3.104	21.566	122913.478	122175.127	-0.604
Aliphatic C9-C12	8508042.000	60.000	3.104	6.370	141800.700	141035.502	-0.543
Aliphatic C12-C16	5470185.000	40.000	6.371	9.754	136754.625	136654.515	-0.073
Aliphatic C16-C21	7891316.000	60.000	9.755	13.108	131521.933	131630.895	0.083
Aliphatic C21-C28	9895762.000	80.000	13.109	16.762	123697.025	122691.322	-0.820