

Continuing Calibration Report for SequenceID : FC022725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068318.D

Aliphatic C28-C40	12764911.000	120.000	16.763	21.566	106374.258	102846.462	-3.430
Aliphatic EPH	44343212.000	360.000	3.104	21.566	123175.589	122175.127	-0.819
Aliphatic C9-C12	8436284.000	60.000	3.104	6.370	140604.733	141035.502	0.305
Aliphatic C12-C16	5417400.000	40.000	6.371	9.754	135435.000	136654.515	0.892
Aliphatic C16-C21	7834817.000	60.000	9.755	13.108	130580.283	131630.895	0.798
Aliphatic C21-C28	9889800.000	80.000	13.109	16.762	123622.500	122691.322	-0.759