

Continuing Calibration Report for SequenceID : FC022825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068341.D

Aliphatic C9-C12	8509686.000	60.000	3.104	6.369	141828.100	141035.502	-0.562
Aliphatic C12-C16	5481187.000	40.000	6.370	9.754	137029.675	136654.515	-0.275
Aliphatic C16-C21	7916038.000	60.000	9.755	13.107	131933.967	131630.895	-0.230
Aliphatic C21-C28	9928727.000	80.000	13.108	16.761	124109.088	122691.322	-1.156
Aliphatic C28-C40	13203681.000	120.000	16.762	21.565	110030.675	102846.462	-6.985
Aliphatic EPH	45039319.000	360.000	3.104	21.565	125109.219	122175.127	-2.402