

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:		
Project:	RFP 905		Date Received:		
Client Sample ID:	Q1421-04MSDMSD		SDG No.:	Q1421	
Lab Sample ID:	Q1421-06MSD		Matrix:	water	
Analytical Method:	NJEPH		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	2000	uL
Soil Aliquot Vol:		uL	Test:	EPH	
Prep Method :	SW3510				

File ID :	Dilution:	Prep Date :	Date Analyzed :	Prep Batch ID
FC068313.D	1	02/26/25	02/27/25	PB166883

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
Aliphatic C9-C12	Aliphatic C9-C12	203		5.76	30.0	ug/l
Aliphatic C12-C16	Aliphatic C12-C16	160		3.61	20.0	ug/l
Aliphatic C16-C21	Aliphatic C16-C21	241		5.14	30.0	ug/l
Aliphatic C21-C28	Aliphatic C21-C28	321		8.97	40.0	ug/l
Aliphatic C28-C40	Aliphatic C28-C40	538		17.0	60.0	ug/l
SURROGATES						
3383-33-2	1-chlorooctadecane (SURR)	34.8		40 - 140	70%	SPK: 50
84-15-1	ortho-Terphenyl (SURR)	0.00		40 - 140	0%	SPK: 50

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID:	Q1421-06MSD	Acq On:	27 Feb 2025 14:40
Client Sample ID:	Q1421-06MSD	Operator:	YP/AJ
Data file:	FC068313.D	Misc:	
Instrument:	FID_C	ALS Vial:	16
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aliphatic C9-C12	3.104	6.370	14340746	101.682	300	ug/ml
Aliphatic C12-C16	6.371	9.754	10947218	80.109	200	ug/ml
Aliphatic C16-C21	9.755	13.108	15865145	120.528	300	ug/ml
Aliphatic C21-C28	13.109	16.762	19693441	160.512	400	ug/ml
Aliphatic C28-C40	16.763	21.566	27692602	269.262	600	ug/ml
Aliphatic EPH	3.104	21.566	88539152	732.092		ug/ml
ortho-Terphenyl (SURR)	0.000	0.000	0	0		ug/ml
1-chlorooctadecane (SURR)	12.841	12.841	4007052	34.84		ug/ml
Aliphatic C9-C28	3.104	16.762	60846550	462.831	1200	ug/ml