

WATER EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Weston Solutions, Inc.
Lab Code: CHEM **Cas No:** Q1421 **SAS No :** Q1421 **SDG No:** Q1421
Sample No : Q1421-05 **Datafile:** FC068312.D
Client ID : P001-CLAY-CF01-01MS

COMPOUND	SPIKE ADDED ug/l	SAMPLE CONCENTRATION ug/l	MS/MSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aliphatic C9-C12	300	0	201	67		(40-140)
Aliphatic C12-C16	200	8.76	158	75		(40-140)
Aliphatic C16-C21	300	0	238	79		(40-140)
Aliphatic C21-C28	400	0	320	80		(40-140)
Aliphatic C28-C40	600	60.6	531	78		(40-140)