

WATER EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Weston Solutions, Inc.
Lab Code: CHEM **Cas No:** Q1421 **SAS No :** Q1421 **SDG No:** Q1421
Sample No : Q1421-05 **Datafile:** FD049121.D
Client ID : P001-CLAY-CF01-01MS

COMPOUND	SPIKE ADDED ug/l	SAMPLE CONCENTRATION ug/l	MS/MSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aromatic C10-C12	200	0	188	94		(40-140)
Aromatic C12-C16	300	12.2	304	97		(40-140)
Aromatic C16-C21	500	43.0	542	100		(40-140)
Aromatic C21-C36	800	0	888	111		(40-140)