

WATER EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Weston Solutions, Inc.
Lab Code: CHEM **Cas No:** Q1421 **SAS No :** Q1421 **SDG No:** Q1421
Sample No : Q1421-02 **Datafile:** FD049142.D
Client ID : P001-CLAY-CF01-01MS

COMPOUND	SPIKE ADDED ug/l	SAMPLE CONCENTRATION ug/l	MS/MSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aromatic C10-C12	200	0.47	207	130		(40-140)
Aromatic C12-C16	300	0.65	364	95		(40-140)
Aromatic C16-C21	500	2.61	755	371	*	(40-140)
Aromatic C21-C36	800	0	1210	151	*	(40-140)