

WATER EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Weston Solutions, Inc.
Lab Code: CHEM **Cas No:** Q1421 **SAS No :** Q1421 **SDG No:** Q1421
Sample No : Q1421-03 **Datafile:** FD049143.D
Client ID : P001-CLAY-CF01-01MSD

COMPOUND	SPIKE ADDED ug/l	SAMPLE CONCENTRATION ug/l	MS/MSD CONCENTRATION ug/l	% REC	Qual	RPD QC LIMITS QC Limit Of RPD
Aromatic C10-C12	200	0.47	197	135		3.77 (40-140) 25
Aromatic C12-C16	300	0.65	347	100		5.13 (40-140) 25
Aromatic C16-C21	500	2.61	721	378	*	1.87 (40-140) 25
Aromatic C21-C36	800	0	1170	146	*	3.37 (40-140) 25