

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Weston Solutions, Inc.
Lab Code: CHEM **Cas No:** Q1421 **SAS No :** Q1421 **SDG No:** Q1421
Sample No : PB166883BS **Datafile:** FC068307.D
Client ID : PB166883BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aliphatic C9-C12	300	195	65		(40-140)
Aliphatic C12-C16	200	179	90		(40-140)
Aliphatic C16-C21	300	306	102		(40-140)
Aliphatic C21-C28	400	416	104		(40-140)
Aliphatic C28-C40	600	688	114		(40-140)