

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Weston Solutions, Inc.
Lab Code: CHEM **Cas No:** Q1421 **SAS No :** Q1421 **SDG No:** Q1421
Sample No : PB166883BS **Datafile:** FD049116.D
Client ID : PB166883BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aromatic C10-C12	200	205	102		(40-140)
Aromatic C12-C16	300	352	117		(40-140)
Aromatic C16-C21	500	679	135		(40-140)
Aromatic C21-C36	800	1070	133		(40-140)