

Continuing Calibration Report for SequenceID : FC022725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068311.D

Aliphatic C9-C12	8524811.000	60.000	3.104	6.370	142080.183	141035.502	-0.741
Aliphatic C12-C16	5481046.000	40.000	6.371	9.754	137026.150	136654.515	-0.272
Aliphatic C16-C21	7930526.000	60.000	9.755	13.108	132175.433	131630.895	-0.414
Aliphatic C21-C28	9943045.000	80.000	13.109	16.762	124288.063	122691.322	-1.301
Aliphatic C28-C40	12574868.000	120.000	16.763	21.566	104790.567	102846.462	-1.890
Aliphatic EPH	44454296.000	360.000	3.104	21.566	123484.156	122175.127	-1.071